Mass Spectrometry in Medical Diagnostics: Principles and Applications

Editors

Karar Abbas Fiai
Medical Device Technologies, Al-Hikma University, Iraq

Mustafa Mohammed Kadhim
Medical Device Technologies, Al-Hikma University, Iraq

Laith Salem Abdul-Hussein
Medical Device Technologies, Al-Hikma University, Iraq

Mustafa Ali Hussein Ali
Medical Device Technologies, Al-Hikma University, Iraq

Bright Sky Publications ™
New Delhi

Published By: *Bright Sky Publications*

Bright Sky Publication
Office No. 3, 1st Floor,
Pocket - H34, SEC-3,
Rohini, Delhi, 110085, India

Editors: *Karar Abbas Fiai, Mustafa Mohammed Kadhim, Laith Salem Abdul-Hussein and Mustafa Ali Hussein Ali*

Edition: *1st*

Publication Year: *2025*

Pages: *88*

Paperback ISBN: *978-93-6233-203-5*

E-Book ISBN: *978-93-6233-145-8*

DOI: *https://doi.org/10.62906/bs.book.339*

Price: *₹ 455*

Contents

Abstract

Mass spectrometry has been established as a powerful tool in chemistry for over 100 years. It has made great strides in the last few decades to conquer the medical field. Mass spectrometry in modern medicine is first and foremost associated with clinical applications, although there are other uses, such as medical research or pharmacology. The popularity of mass spectrometry comes from its unique properties of high sensitivity and the ability to identify mixtures of components with a very high level of selectivity. In a clinical context, selective and sensitive analytical detection methods are used. The main areas of mass spectrometry applications are the analysis of low-molecular-weight compounds, biopolymers and proteomics. The essence of mass spectrometric measurements is the determination of the mass-to-charge ratio of ions generated from the sample. Based on a series of mass spectrometric measurements, a mass spectrometric experiment provides a unique signal related to the elemental and isotopic composition of the analytes.

A major factor contributing to the high sensitivity of mass spectrometric measurements is the use of a vacuum as the environment for ions within the mass analyzer region. The ions to be analyzed in a mass spectrometer are guided through the analyzer region, which is usually under high vacuum to avoid collision with molecules or gas atoms, resulting in a reduction in the intensity of the ion signal of the analyzed sample.

Chapter - 1

Introduction to Mass Spectrometry in Medical Diagnostics

Mass Spectrometry-based analyses have undeniably demonstrated a pivotal and transformative role in the expansive and continually evolving field of medical diagnostics. The inception of mass spectrometry can be traced back nearly a century, with its development experiencing a significant and noteworthy acceleration over the course of the last few decades leading up to the present day. As one of the most vital and essential analytical tools in the realm of scientific exploration and inquiry, mass spectrometry is not only extensively utilized in foundational and basic scientific research, but it is also rapidly and increasingly being applied across numerous critical and diverse areas including, but not limited to, drug discovery, food safety, environmental protection, material science, anti-terrorism efforts, and many others beyond those typically considered. Within these varied and diverse applications, mass spectrometry has grown into a highly specialized branch of study, which is known as proteomics [1, 2, 3, 4, 5, 6].

Among the multitude of applications of mass spectrometry, its increasing utility and effectiveness in the ever-expanding realm of medical diagnostics has garnered considerable and intensified attention from various stakeholders, including researchers, clinicians, and even patients alike. Recent reports and literature have primarily focused on the ongoing development of innovative ambient ionization techniques along with miniature mass spectrometry systems that are specifically aimed at enhancing disease diagnosis and aiding in therapeutic monitoring in real-time conditions. Although there are a number of evident and recognized challenges and limitations associated with the practical implementation and operational use of mass spectrometry in clinical environments, there has been a remarkable and rapid evolution in a multitude of disciplines accompanied by a notable surge in innovative strategies that work diligently toward overcoming these hurdles [7, 8, 2, 9]. This ongoing progress has led to the creation of several direct analytical methods that can be effectively and seamlessly integrated with compact mass spectrometry systems, rendering them particularly suitable for point-of-care applications that are increasingly in demand in today's healthcare landscape. Looking ahead into the future, it is entirely reasonable and justifiable to predict that mass spectrometry will increasingly assume a more central and direct role within the expansive field of biomedicine and its numerous sub-disciplines. With the continued and steadfast advancement of mass spectrometry technologies and methodologies, it is highly likely that common and prevalent diseases will be managed with greater efficacy and precision in the future as a resulting consequence of these innovations. The historical trajectory of mass spectrometry's application in medical diagnostics has been lengthy, nuanced, and, at times, inconsistent, reflecting a complex and multifaceted evolution that mirrors the intricacies of medical progress itself [10, 11, 12, 13, 14, 15].

In a historical context, the early devices used, including the pioneering gas chromatography/mass spectrometry (GC/MS) systems developed back in the day, were often characterized as being large, stationary, and cumbersome instruments that systematically lacked the sophistication, efficiency, and portability of the advanced and highly specialized technologies available today. Despite the limitations and constraints of these initial tools, mass spectrometry has remarkably established itself as a cornerstone technology in the field, flourishing in its application scope and markedly enhancing its technical capabilities over time in various settings. Relatively simple systems, built upon incremental technological advancements and refinements over the years, have matured and evolved significantly to give rise to the sophisticated machines that are now prevalent and widely utilized in laboratories around the world today [16, 17, 18, 19, 20, 21, 22]. On a fundamental level, mass spectrometry instruments have achieved a near-ubiquitous presence within a truly wide range of pharmaceutical environments, where they are relied upon extensively across an impressive array of advanced techniques to deeply analyze molecular weights, intricate structures, and various critical physical properties of numerous chemical compounds and complex mixtures alike. The global market dedicated to the field of mass spectrometry has proven to be immensely large and lucrative, with careful estimates suggesting that it exceeds an astonishing four billion dollars on an annual basis, thereby affirming its undeniable significance and robust economic impact across various sectors and applications. Beyond the traditional confines of pharmaceutical laboratories and various academic research environments, mass spectrometry has successfully established a highly prominent presence and has found significant roles across a wide variety of other industries, such as environmental analysis, food safety measures, and clinical diagnostics. This further demonstrates its tremendous versatility and enduring importance in contemporary diagnostics and analytical practices across the globe, which ultimately benefits countless applications and assists in contributing to continuous advancements in science and technology. Its adoption reflects a commitment to excellence and a response to the ever-evolving needs of industries that rely on precise and accurate analytical methods for quality assurance and compliance with regulatory standards [23, 24, 25, 26, 27, 2, 3, 28, 29, 30, 31].

Chapter - 2

Fundamental Principles of Mass Spectrometry

Mass spectrometry is an exceptionally advanced and intricate analytical technique that is widely and extensively utilized across numerous scientific disciplines and fields. Its primary aim is to precisely determine and measure the mass-to-charge ratio of ions with remarkable precision and accuracy. The foundational principles that underpin mass spectrometry are not merely essential; they have ultimately become fundamental in the broad and expansive field of analytical chemistry, where these principles are often informally referred to as the ABCs of mass spectrometry. In this insightful context, the letter A typically stands for Atomic Weights, which are absolutely critical in the meticulous calculation and determination of the masses associated with different ions, thereby serving as a cornerstone for understanding and elucidating molecular structures [32, 33, 1, 34, 35, 36, 37].

The letter B denotes the mass-to-charge ratio, which represents a crucial measurement that is carefully determined within the confines of a powerful magnetic field, thus enabling researchers, scientists, and analysts to effectively distinguish and differentiate between ions of varying and diverse masses. Furthermore, the letter C encapsulates the sophisticated and complex concept of high-resolution mass spectrometry of Continua, which yields profound, intricate, and detailed insights into the molecular composition of complex mixtures, facilitating advanced and highly capable analytical methodologies. Notably, the first documented description of the term "mass spectrograph," an innovative instrument that has played a significant and transformative role in the ongoing evolution of this analytical technique, can be traced back to the year 1912. This marked a pivotal and historic milestone in the field of analytical science, laying the necessary groundwork for all future advancements in the practice. Within the intricate and complex realm of mass spectrometry, ions are systematically generated from the various molecules of the analyte that is under thorough examination, thus enabling a meticulous and detailed understanding of their intrinsic properties and characteristics. Initially, the sample undergoes a fundamental and pivotal step known as atomization, wherein the solid or liquid substance is carefully transformed into individual atomic particles, all to facilitate a more comprehensive and thorough analysis. Following this foundational and crucial step, a subsequent and essential process termed ionization is employed, during which the produced atoms undergo a specialized electrical charging procedure, ultimately resulting in the generation of either positive or negative ions. This elaborate, comprehensive, and detailed procedure is not only vital for the successful analysis of samples but also plays a critical and indispensable role in the reliable identification and characterization of complex substances encountered across a wide array of various applications. These applications range from rigorous and stringent environmental testing to the sophisticated development of various pharmaceuticals, demonstrating the broad relevance, necessity, and importance of mass spectrometry in modern science and advanced technology [38, 32, 33, 39, 36, 40, 2, 41, 42].

The charged particles that are present in mass spectrometry can be effectively and efficiently separated in space based on their distinct and unique mass-to-charge ratio (m/z), which is a fundamental principle in this analytical technique. This sophisticated and nuanced separation is achieved through the meticulous utilization of an electric field, which plays a crucial and integral role in this analytical process, acting as a driving force in the segregation of ions based on their electrical characteristics. This method of separation is further complemented by the discernible distinctions made through momentum management via a magnetic field, which adds an additional layer of complexity and accuracy to the analysis. Alternatively, one can adopt a different strategy or method, such as time-of-flight analysis, which involves measuring the time it takes for ions to travel a known distance, providing another effective means to accomplish this vital task of ion separation. Ultimately, the current of each individual ion is carefully and meticulously measured with great attention to detail, ensuring that every aspect of the analysis is conducted with precision. This meticulous measurement facilitates the precise and accurate construction of the required and comprehensive mass spectrum that is essential for in-depth analysis of the analyte [43, 44, 45, 46, 47, 48, 49]. All of these vital and essential processes are performed within a meticulously controlled high vacuum environment, designed to eliminate interference from air molecules and other contaminants, thereby guaranteeing utmost accuracy and precision in all measurements undertaken. This advanced and sophisticated analytical technique empowers scientists and researchers to not only determine the elemental composition of an analyte but also to unveil the isotopic ratios of the various elements that constitute it. Moreover, it simultaneously allows for the identification of concentration levels and the presence of multiple components within a single sample, even those that may exist in significantly low quantities or abundances, thus proving remarkably valuable in a variety of scientific investigations. Mass spectrometry (MS) stands out as the quintessential and most reliable method among all analytical techniques in the scientific community, establishing itself firmly as a cornerstone and a fundamental pillar in the realm of analytical analysis. As a broad umbrella term, mass spectrometry encompasses a diverse array of technologies, methodologies, and approaches that meticulously unveil the intricate and complex nature of analytes, derived from the precise and methodical measurement of their mass-to-charge ratio [43, 45, 44, 46, 50, 51, 49, 52, 53].

A mass spectrometric setup is composed of three fundamental and essential components: an ion source, which initiates the analytical process by generating charged ions from the sample; an analyzer, which serves the crucial function of separating these ions based on their specific m/z ratio and directing them for detection; and a detector that plays a vital role by generating an electronic signal that is directly proportional to the abundance of ions corresponding to the specific m/z ratio being considered in the analysis. From these three key components-ionization process, mass analysis stage, and detection-a comprehensive and thoroughly rigorous approach is undertaken to thoroughly address the various instrumental technologies that have been meticulously developed and are currently in use throughout the historical evolution of the mass spectrometry field. This evolution has seen significant advancements, leading to the development of various techniques tailored to meet specific analytical needs across different fields of research and application. As an essential introductory step, it is crucial to conduct a thorough and detailed examination of the principles that underlie the construction of stable, reliable, and useful ions, as well as the diverse range of innovative techniques that are employed to efficiently separate them for in-depth analysis. A detailed exploration is also undertaken to determine which mass analysis techniques, categorized under this broad and generic term, are currently available or have been developed over the years, including the rationales and scientific reasoning behind their development and implementation. The initial chapter serves as an enlightening introduction to the expansive, dynamic, and evolving field of mass spectrometry, covering its fundamental principles and aiming to educate both those who are completely unfamiliar with the field as well as those who may possess a general understanding but are seeking a more detailed focus on the better-known and widely used techniques essential for practical applications. The comprehension of these foundational principles and the significance of adhering to them greatly influences the understanding and ability to successfully conduct mass analysis in various scientific contexts, enabling researchers and scientists to further their investigations with enhanced clarity, accuracy, and precision, ultimately advancing the frontiers of scientific knowledge and discovery [34, 32, 54, 55, 42, 56, 57].

In the expansive and intricate realm of clinical diagnosis, mass spectrometry (MS) methods have undeniably carved out a significant and vital niche, establishing an essential and indispensable role in the evolving and dynamic landscape of modern medical science and practice. This importance is underscored by a broad and varied spectrum of applications that span both targeted and untargeted analyses, as demonstrated by an extensive and diverse array of scientific studies that consistently illustrate its profound and multifaceted utility. As an exceptionally versatile and powerful instrument for targeted analysis, mass spectrometry is particularly adept at closely monitoring the degradation products of specific pharmaceutical drugs or determining the precise and exact concentrations of a target drug that may be present in various biological fluids, such as blood, urine, or tissues. This remarkable capability is crucial for conducting essential toxicological assessments and ensuring thorough and comprehensive safety evaluations in various clinical contexts, particularly when it comes to making informed and evidence-based decisions that significantly impact patient care and management outcomes [58, 59, 60, 4, 61].

Conversely, the high sensitivity and remarkable specificity that are inherent in MS practices have also firmly established it as an exceptionally powerful tool for conducting detailed and comprehensive research within a diverse array of biological samples. This enabling capacity allows researchers to achieve unprecedented and unparalleled insights into various health conditions, fostering much deeper understandings of the intricate disease mechanisms that underpin numerous significant medical issues. Indeed, a number of groundbreaking studies have remarkably illustrated that mass spectrometry possesses the extraordinary and unique capability to detect cancerous or precancerous states in serum samples long before traditional clinical diagnosis methods can even aspire to achieve similar results. The extraordinary performance of mass spectrometry in this critical area is a compelling and noteworthy feature of its application in the field of oncology, marking a crucial advancement in diagnostic techniques and therapeutic strategies [27, 1, 62, 63].

This truly remarkable ability to effectively and accurately identify malignancies at an earlier and more treatable stage holds immense potential to greatly facilitate timely and life-saving specific treatment interventions for patients. By identifying these malignancies sooner, it significantly improves prognosis and dramatically increases survival rates, which is a primary and crucial goal in modern healthcare practices. Likewise, the innovative and untargeted analysis of small molecules that are present in urine has been effectively employed not only to diagnose kidney cancer but also to detect a wide range of other serious medical conditions, such as various complex metabolic disorders and cardiovascular diseases. This noteworthy capability underscores the extensive versatility and broad applicability of this transformative technology across a multitude of clinical scenarios and circumstances. It greatly illustrates its remarkable capability to address diverse healthcare challenges that may arise throughout patient care, making it an invaluable tool in the ongoing and concerted efforts to enhance diagnostic accuracy and treatment effectiveness in contemporary medical practice [64, 65, 58, 62, 1, 66].

Collectively, these pivotal studies strongly underscore the immense potential of mass spectrometry-based metabolomics methodologies for enabling non-invasive early screening and accurate diagnosis of various types of cancer. Such significant advancements ultimately provide the medical community with essential and innovative tools that empower them to offer more precise and highly personalized treatment strategies and approaches, which are carefully tailored to meet the unique needs of individual patients. This advancement not only presents the promising possibility of improved stratification of treatment plans but also allows for the effective and continuous monitoring of disease evolution over time. This ongoing vigilance is crucial for promptly adapting treatment strategies as necessary based on the patients' responses to therapies and their evolving health status. The thoughtful integration of these advanced methodologies can significantly enhance patient outcomes by ensuring that each individual receives the most suitable intervention at precisely the right time, thereby maximizing the overall efficacy of treatment and improving survival rates across diverse patient populations. Through these remarkable advances, mass spectrometry is strongly poised to revolutionize the landscape of clinical diagnostics dramatically, ultimately contributing to enhanced patient outcomes and a deeper understanding of complex biological processes that govern human health. This revolutionary change is paving the way for innovative healthcare solutions that were previously unimaginable, setting the stage for a new era in medical diagnostics and patient care that aligns closely with the advancements in technology and scientific discovery. As our understanding of molecular biology continues to expand, the role of mass spectrometry in clinical practice will undoubtedly grow, leading to more effective diagnosis, treatment, and management strategies for a wide array of diseases and health conditions [23, 67, 1, 27, 68, 16, 69, 3, 70, 71, 72].

2.1 Mass Analyzer Types

Along with the myriad of features that can be found in advanced mass spectrometers, it is important to recognize that the fundamental nature of their underlying physical principles undoubtedly plays a decisive and crucial role in determining, not only the practical utility of the equipment itself, but also the specific application context of various analytical methods that are utilized in scientific research laboratories. Mass analyzers, which represent a vital subsection of mass spectrometers, are entrusted with the critical and complex task of determining the specific mechanisms by which ions of diverse masses are methodically separated from one another. Consequently, this methodical separation dictates how the subsequent detection process operates, enabling the detector to accurately record both the presence and the distinct identity of these ions of interest. These analyzers are generally categorized based on their operational principles as a primary factor and not solely on their specific analytical methods for separation and identification. As such, they can be classified into two broad categories that encompass a diverse array of technologies and techniques employed in the analytical field, each with its own methodological nuances and specialized applications to suit various research needs [73, 43, 45, 74, 75].

Mass selection and analysis that strategically relies on the motion of ions within a well-defined combination of both radio-frequency fields and direct current electric fields has led to the development of three main types of mass analyzers: Quadrupole (Q) analyzers, Ion Trap analyzers, and Fourier Transform based (FT-MS) analyzers. Within the framework of each of these sophisticated instruments, only those ions that possess a prescribed mass-to-charge ratio, denoted as m/z, are capable of completing stable trajectories. This characteristic directly influences several important operational parameters, including mass resolution as well as the bandwidth of the analyte ions detected by the system. These particular types of mass analyzers rank among the most commonly employed and utilized in various analytical applications across a broad spectrum of scientific disciplines. Furthermore, due to their relatively high levels of analytical performance, which is expertly balanced against straightforward operational characteristics, they are frequently regarded as the preferred choice for commercially available mass spectrometry systems. These systems have been especially designed for practical use in diagnostic laboratories and other scientific research environments where precision and accuracy are paramount [38, 73, 76, 77, 78, 79, 80, 75, 81, 82, 83].

Alternatively, there exist several types of mass spectrometers that operate in a distinctive manner, specifically one in which ions are subjected to a uniform acceleration that intriguingly remains independent of their mass-to-charge ratio (m/z). This unique and sophisticated process results in the kinetic energy that is acquired during the acceleration being effectively channeled into velocity, which is then utilized as a crucial mechanism to separate distinct ion species that possess varying m/z values. This fascinating principle is prominently exemplified in the widely recognized and utilized Time-of-Flight (TOF) analyzer. The fundamentally different operational principles of these various instruments give rise to analytical methods that feature an array of distinct characteristics, which are generally designed and tailored carefully for specialized applications that cater to specific analytical tasks across a variety of fields. However, it is certainly worth noting that each of these methods has its own individual advantages- for instance, the TOF instrument provides outstanding resolution coupled with remarkable speed; Orbitrap-type instruments significantly enhance the stability of Fourier Transform Mass Spectrometry (FT-MS), ensuring consistently high-quality results; furthermore, ion trap-based analyzers offer sensitive detection that is characterized by a 'default' unit mass resolution, which facilitates the efficient analysis of complex mixtures through advanced and sophisticated mass filtering techniques. With ongoing and rapid technological advancements in this dynamic field, these methods are continuously evolving and being refined, making it an expanding and progressive discipline that consistently pushes the boundaries and challenges the conventional limitations typically associated with other mass analyzers. This evolution in technology invariably opens the door for a multitude of exciting new applications and innovative solutions that were previously unimagined. Attaining a comprehensive understanding of the principles underlying various mass analyzers is of utmost importance, as this knowledge is essential to effectively utilize these sophisticated and highly advanced instruments to their fullest potential, maximizing their capabilities and applications in analytical chemistry and beyond, fostering new discoveries [84, 85, 86, 87, 88, 89, 90, 91, 92, 93, 94].

2.2 Ionization Techniques

The key fundamental aspect to fully understanding the intricate and sophisticated process of mass spectrometry lies in the crucial and critical generation of ions from a given analyte substance. During the elaborate and multifaceted course of the ionization process, the analyte or its corresponding solution undergoes a significant transformation and is thoroughly converted into a comprehensive set of gaseous ions. For instance, Electrospray Ionization (ESI) is a complex technique that, unlike some other methods, does not represent a direct ionization process as clearly observed in techniques such as Atmospheric Pressure Chemical Ionization (APCI) or Electron Ionization (EI). Instead, it involves an infusion process during which the precise act of ionization enables the efficient evaporation of ions along with the solvents and other components that are present during this multifaceted process. The ions that are evaporated undergo notable and significant changes at the liquid surface leading to the establishment of a Chapman cycle during this delicate and complex dance of molecules. This particular method is especially suitable for applications in critical drug testing and various essential sanitation assessments, providing consistently reliable results. Due to its unique ability to maintain no direct contact with the object being analyzed, many ionization sources first transition the sample into a gaseous phase before any subsequent analysis. During the complex ionization process, the analyte, whether it is in liquid, solid, or gaseous form, typically contains a significant and influential proportion of solvent or matrix material that plays a key role in the overall ionization process. As the ionization continues, a carefully electrosprayed droplet experiences both extensive evaporation followed by a dramatic explosion, resulting in CRIME (Chemical Reaction in Mass Spectrometry) products that are subsequently analyzed through various mass spectrometry techniques. In stark contrast, ions that are formed via Electron Ionization Mass Spectrometry (EI-MS) are generated through a direct and forceful ionization approach, precisely targeting the solid or liquid sample while it is already situated in the gas phase. This distinct methodology thus provides a different and unique set of capabilities in the analysis of complex mixtures and compounds, which are paramount in advancing the fields of analytical chemistry and biochemistry significantly [95, 96, 97, 98, 99, 100, 101, 102, 103, 104].

Ionization of the sample stands out as one of the most crucial and significantly important parts of mass spectrometric analysis, as it plays an essential role in the determination and quantification of an extensive range of various compounds. The ionization process must be executed with a high degree of universality and efficiency simultaneously. This is vital, as it enables the effective ionization of a broad and diverse range of compounds that are present in significant concentrations, which is pivotal for acquiring reliable and reproducible data. The meticulous selection of the ionization method-whether it be Electrospray Ionization (ESI), Atmospheric Pressure Chemical

Ionization (APCI), Atmospheric Pressure Photon Ionization (APPI), Desorption Ionization on a Chip (DIOS), Matrix-Assisted Laser Desorption/Ionization (MALDI), and many other innovative techniques that are available-can be influenced by the unique physical and chemical properties inherent to the samples themselves. In some instances, the method options available may also be dictated by the specific type of mass spectrometer that is being employed in the analysis process. Consequently, the direct ionization of samples, particularly those collected from insect traps, is quite rare and often not commonly discussed in the existing literature regarding mass spectrometry. Furthermore, various modifications and improvements of the Atmospheric Pressure Ionization (API) source, as well as other types of ion sources, have been successfully accomplished and refined by numerous suppliers over time to enhance their efficiency. There are also ongoing reports that highlight the emergence of open access ion sources, which provide alternative options and methodologies for researchers within the scientific community to explore. In order to thoroughly understand the comprehensive effect of the ion source on the results obtained from mass spectrometric analysis, alongside the overall capability of mass spectrometry from a broader perspective, a detailed and extensive overview of the underlying principles of the various ionization methods currently employed in the field is strongly advised and highly recommended for those seeking to deepen their knowledge and expertise in this essential area of analytical chemistry [105, 106, 36, 107, 108, 73, 109, 2].

Chapter - 3

Instrumentation in Mass Spectrometry

Medical diagnostics are progressively relying more on mass spectrometry (MS), which is becoming an indispensable and increasingly influential technique within the field. To genuinely comprehend the robust, varied, and extensive nature of the information provided by this advanced analytical method, it is crucial to delve into some fundamental and foundational concepts regarding mass spectrometry. This discussion will encompass a concise yet enlightening overview of the different key components that constitute a typical MS setup, the specific and vital steps involved in meticulously preparing samples for thorough analytical evaluation, as well as the remarkable diversity of analyzable substances, which range from small organic molecules to large complex biomolecules, that can be thoroughly studied with this powerful and versatile technique. Understanding these critical elements will not only illuminate the myriad advantages of mass spectrometry but will also demonstrate its extensive applications in a diagnostic framework, emphasizing its transformative role in improving medical practices, leading to better patient outcomes and more accurate diagnoses. The integration of mass spectrometry into routine diagnostics represents a significant advancement in the way we understand and address medical conditions, guiding clinicians in making more informed decisions based on precise, reliable data derived from the sophisticated analytical capabilities of MS technology [24, 1, 25, 4, 10, 2, 3, 60, 110].

The key components that characterize the remarkable and extensive variety of mass spectrometry (MS) setups will be dealt with in great detail, offering insights into their diverse configurations and methodologies. Following this, the intricate analytical process involved in these setups will be discussed extensively, which encompasses the essential and critical steps that occur systematically between the initial sample (S) and the final observation (O). These fundamental steps, which are vital for accurate analysis, include the careful transportation of the sample, the pivotal process of ionization (I), the crucial separation step (S), and the vital monitoring or detection (D) phase. It will be elaborated on how these numerous steps significantly affect one another, highlighting their interdependencies, and how they need to align perfectly in order to obtain reliable and reproducible analytical results that are critical for scientific investigations across various disciplines. Moreover, the importance of diligently maintaining the quality and integrity of each step in this analytical chain, as well as the rigorous calibration processes that are necessary to ensure accuracy throughout the entire procedure, will also be addressed comprehensively. The various techniques and innovative technologies that have pioneered the use of mass spectrometry within the realm of medicine shall be briefly discussed, as they provide an essential backdrop to the relatively recent and innovative application of these methodologies in disease diagnostics and management. On the side of early medical diagnostics, notable historical examples will include Isotope Ratio Mass Spectrometry (IRMS), which was notably utilized for detecting prohibited doping substances with great precision during the infamous 1996 Tour de France, and gas chromatography-mass spectrometry (GC-MS), which was used effectively for the detection of cocaine in urine samples as early as 1973, demonstrating the long-standing relevance of these techniques. Furthermore, it is intended to provide a thorough overview of significant and important recent advancements in mass spectrometry that are particularly relevant for the growing involvement in many critical areas of clinical diagnostics. These advancements will serve as prominent examples in the subsequent articles that follow, showcasing their potential applications. These developments encompass, among other things, the rise of a new commercial sector that provides an increasingly diverse range of methods and techniques aimed at ensuring broad accessibility of mass spectrometry technologies outside of specialized institutions and traditional laboratories; the ongoing miniaturization of instruments, which allows for more compact and portable setups to facilitate their use in various environments; and significant advances in detection technologies that empower high-throughput analysis of complex biological substances across various clinical applications, ultimately enhancing the speed and efficacy of diagnostic processes for a variety of medical conditions [38, 1, 24, 27, 10, 111, 6, 112, 113].

The selection of the analyzing instrumentation plays an undeniably crucial and significant role in shaping the multitude of inquiries that can be posed regarding various samples under investigation in today's rapidly evolving and advancing scientific milieu. Within the vast and expansive realm of medical research, there is a continually expanding and greatly varied array of substances being examined, which encompasses both endogenic compounds such as proteins, lipids, hormones, and antibodies, alongside a diverse assortment of various medicinal products. These diverse substances frequently occur within exceedingly intricate and complex mixtures that encompass a wide range of concentrations and distinct characteristics, spanning several orders of magnitude and a multitude of variations. This inherent complexity necessitates the utilization of a truly diverse array of analytical methods, although only a select few of these sophisticated techniques can be effectively conducted on a given mass spectrometry (MS) setup without encountering significant operational and analytical challenges. Alongside these crucial considerations, it is often essential to pose at least some initial inquiries regarding a substance's quantitative characteristics, including its mass, empirical formula, and other related parameters, which may require the conducting of multiple detailed acquisitions utilizing a variety of different analytical techniques and sophisticated instruments. Conversely, when faced with particularly challenging and complex conditions-such as difficult sampling situations, low concentrations of essential substances, or the presence of unstable, volatile, and reactive compounds-alternative methodologies become absolutely essential for a comprehensive, in-depth analysis. From a strategic perspective, the application of mass spectrometry finds significant and widespread application across various therapeutic fields, particularly in the thorough and systematic assessment of a wide range of pharmaceuticals and their metabolic by-products. It plays a pivotal and critical role in accurately quantifying drug levels, understanding the nature and extent of metabolic processes, and thoroughly evaluating potential drug-drug interactions that may occur between different substances. The insights gleaned from such meticulous and rigorous analyses can play an invaluable and essential role in safeguarding against inappropriate and potentially dangerous dosages, as overdosing on certain substances can result in a plethora of serious and potentially life-threatening adverse effects, which can manifest even at relatively low concentrations in the bloodstream and require immediate attention. In this detailed discussion, we will delve into a comprehensive evaluation of the potential of mass spectrometry diagnostics, particularly within the vital and essential context of toxicological research. It is important to note that this thorough analysis does not intend to delve into the myriad of other diverse applications where mass spectrometry has firmly established its significant foothold, providing transformative insights across various scientific fields and disciplines. Moreover, it is worth mentioning that there has been relatively minimal focus on the underlying theoretical frameworks and principles that govern mass spectrometry as a scientific discipline, which can be absolutely essential in understanding the full spectrum of its capabilities and numerous applications in contemporary research and investigation [23, 73, 114, 41, 115, 1, 35, 116, 117, 118, 119].

Given the specified terms and recognizing that medical treatments and curative procedures can differ significantly based on the specific substance that has been ingested, it is considered adequate to refer to the various drugs involved without delving into exhaustive or overly detailed explanations. This approach is aimed at maintaining clarity while also acknowledging the complexity and variability that are inherent in the different substances and their associated remedies. By adopting this method, we ensure that the focus remains clear and concise, allowing for a better understanding without overwhelming the reader with excessive detail that might confuse rather than inform [120].

3.1 Ion Sources

Ion sources play an absolutely crucial role in the expansive and intricate mass spectrometry (MS) process, serving as essential and pivotal components that conveniently facilitate the conversion of neutral samples into ions of a specific nature. These critical parts of a mass spectrometer function as dedicated and specialized mechanisms focused solely on producing the necessary ions that are required for thorough and precise analysis. Within the vast and diverse realm of mass spectrometry, there are recognized three primary types of ionization techniques, widely acknowledged for their remarkable effectiveness and utility: electron impact (EI), chemical ionization (CI), and fast atom bombardment (FAB). Furthermore, in addition to these commonly utilized techniques, there exist several other, less frequently employed types of ionization methods, including field desorption, thermospray, spark ionization, and direct liquid introduction. Each of these methods presents unique advantages and is tailored to different kinds of analytical needs, adding to the rich toolkit available to researchers and scientists. Among the various techniques available, the first two-electron impact and chemical ionization-will be explored and examined in greater detail, due to their significant importance and wide applicability, particularly in the critical and evolving field of medical diagnostics. Their diverse and extensive uses extend to numerous applications across various sectors, showcasing the remarkable versatility and critical nature of effective ionization in the analytical process. This versatility highlights the significance of ion sources in advancing our understanding of complex biological samples and chemical compounds. It emphasizes why a proper and comprehensive understanding of these vital ionization methods is essential for achieving optimal results in the sophisticated art of mass spectrometry, ensuring that analyses are both accurate and highly informative. Moreover, the dynamic capabilities of various ionization techniques allow scientists to explore and unearth new information from even the most challenging samples, thus further underscoring the crucial need for ongoing research and continuous development in this critical area of scientific inquiry. The importance of these techniques cannot be overstated, given that advances in ionization serve as foundational building blocks upon which the entire mass spectrometry framework rests. This progress pushes forward the boundaries of what is known in the fields of chemistry, biology, and beyond. Ultimately, it contributes to groundbreaking discoveries and innovative technologies that shape our understanding of the world around us. By harnessing the power of these ion sources and continually refining the techniques involved, researchers can unlock new insights that drive scientific progress and improve diagnostic capabilities in health and medicine, thus emphasizing the ever-growing relevance of ionization techniques in contemporary analytical chemistry [95, 121, 122, 123, 124, 125, 57, 109, 55, 126].

In Electron Impact (EI), sample molecules undergo a complex process of ionization that occurs due to the generation of highly energetic electrons produced within a carefully predefined range of 70 eV. These energetic electrons are generated by a filament specifically designed to bombard the entrance capillary of an ion source, which is crucial for the proper functioning of the ionization process. This entrance capillary plays an essential role as it encloses the entire ionizer assembly, ensuring that the electron beam is effectively confined within its bounds to prevent any unwanted arcing that could significantly interfere with the critical ionization process. The energetic electrons collide with sample molecules, leading to the formation of M-1 radical cations and releasing an additional electron (e-) for each ion that is created in this process. The ions generated through this method are then carefully focused and directed toward the Quadrupole mass filter. This is accomplished through the application of the Repeller DC potential, which is instrumental in accurately guiding these ions so they can be further analyzed. In contrast, Chemical Ionization (CI) employs a markedly different methodology that uses a less energetic and more gentle process to produce ions. In the CI approach, sample molecules are transported within an inert gas and subjected to bombardment from a medium-energy electron beam that operates at a more moderate energy level of around 35 eV. This interaction results in a charge exchange occurring during the collision events, ultimately leading to the formation of a reagent ion alongside the corresponding sample ion. To enhance the quality and resolution of the spectral width in the analysis, a Q0 air lens is deployed strategically to focus attention on both the surplus reagent ions as well as the ions that are generated from the sample itself.

Furthermore, there are a variety of sophisticated strategies and techniques involving specific chemicals that can be utilized in conjunction with the mass filter to further refine and optimize this process. The ions that are created with a positive charge, known as radical cations, are drawn in through an entrance cone, where they undergo focusing by a lens into what is technically referred to as an ion guide during the RF-only operational phase. The ion guide, which may also be referred to as a quadrupole ion deflection cell (Q0), serves the purpose of facilitating modifications in the trajectory of these ions, effectively partitioning them onto a central electrode that plays a critical role in their analysis. This central electrode can be dipped into the collision cell, allowing it to further dissect and explore the complexities of the ionization process. During the interaction phase, ions possess the capability to excite anions that are generated through a 19 eV negative chemical ionization source, enriching the overall dataset. During the collision events that occur, both positive and negative ions may experience energy loss, resulting in the release of any unreacted ions and neutral molecules that remain post-interaction. Typically, levels of disease biomarkers are detected in the nanogram per deciliter (ng/dL) range, highlighting the crucial importance of ion sources within the context of analytical chemistry, as they are fundamentally responsible for generating the ions necessary for further analysis. Therefore, a comprehensive understanding of the underlying mechanisms of ion production is essential for the optimization of mass spectrometry (MS) instruments and the associated analytical protocols that govern their operation. It is imperative to choose the most suitable ion source in relation to the specific samples being analyzed, while also ensuring the maintenance of an apparatus that operates both efficiently and stably. The EOF was a preparatory window that capitalized on the numerous capabilities of EI, which is known for its impressive inter-laboratory reproducibility. This reproducibility greatly enhances the reliability and consistency in results obtained through various analyses. For critical medical analyses involving blood, serum, or plasma samples, EI presents a highly valuable methodological approach, and the recent advancements in the field have the potential to significantly reduce run times, possibly by as much as half, which would greatly streamline the workflow in laboratories. These developments could lead to even greater efficiencies and improved outcomes in the analysis of clinical samples [127, 128, 129, 130, 131, 132, 133].

3.2 Mass Analyzers

One of the essential components of a mass spectrometry setup is undoubtedly the mass analyzer, which plays a pivotal role in the overall functionality of the entire analytical system. This crucial instrument is responsible for the precise separation of ions, specifically based on their unique mass-to-charge ratios, a fundamental concept that is essential in facilitating an in-depth understanding of the composition of various samples across different domains of scientific inquiry. Over the past two decades, various mass analyzer techniques have undergone rapid advancements and substantial developments, creating significant contributions that have remarkably enhanced the sensitivity and specificity of mass spectrometry, establishing it as an invaluable tool in many diverse scientific fields. A wide variety of mass analyzers have been developed and refined during this period, including but not limited to quadrupole analyzers, ion trap systems, advanced time-of-flight (TOF) instruments, linear ion trap devices, state-of-the-art orbitraps, and Fourier transform-based analyzers. Each of these analyzers possesses its own distinct set of advantages along with inherent limitations that must be taken into careful consideration in the context of their application. The various trade-offs between important factors such as mass resolution, mass accuracy, dynamic range, and the overall complexity of their operation need to be thoroughly examined when selecting the most appropriate analyzer for a particular application, specific research goals, and desired experimental outcomes. Understanding these critical factors is essential for researchers striving to obtain high-quality data and optimize their experimental results effectively in their respective studies. Properly navigating these considerations can lead to significant improvements in the quality and reliability of data generated through mass spectrometry techniques, ultimately advancing scientific knowledge and applications in various areas [84, 73, 74, 134, 1, 94, 122, 135, 124, 136, 5 . 5]

In the previous century, the quadrupole mass filter emerged as one of the most commonly and widely utilized devices, particularly when used alongside a magnetic sector analyzer for various analytical purposes. The ongoing decision-making process regarding the selection of a quadrupole mass analyzer is complex and multifaceted, especially when it is combined with other types of mass spectrometers, necessitating careful and continuous consideration of numerous critical issues. These issues are intrinsically related to the design and development of high-performance research instrumentation in contemporary analytical chemistry. In making this choice, one must take into account a wide range of practical factors, which include the specific intended use of the system in question, the selection process that is involved with other essential mass spectrometric components, the degree of automation through computer control of the entire system, along with cost considerations-thus presenting a complex trade-off between differing performance metrics, simplicity of operation, overall capital costs, and the commercial availability of the necessary equipment. With the recent and significant advances in technology, various types of ion trap mass analyzers have been thoroughly examined and reflected upon, with the aim of creating mass spectrometers equipped with the most versatile capabilities and the highest performance standards achievable in the field. The development of highly sensitive and accurate mass determination capabilities has greatly enhanced the ability to allow for the unambiguous detection of analytes-substances of interest-and has promoted high-quality structure elucidation of these various substances, even when present at low concentrations. Because of this enhancement, a quadrupole mass filter, especially when coupled to a time-of-flight analyzer, can achieve exceptionally high levels of mass accuracy (>10 ppm). This capability significantly broadens the scope and range of analyzable compounds, particularly within complex matrices commonly encountered in research. Mass accuracy has been a particular point of interest and focus in research and development throughout recent years, and it has been widely applied in both precise mass determination as well as in tandem mass spectrometry applications. This noteworthy development in the design and application of mass analyzers, combined with substantial advancements in related data analysis software, has contributed significantly to several important areas of application in analytical chemistry. These advancements include the ability to effectively identify one or multiple known unknowns, as well as confirming or refuting qualitative results obtained from various experimental methodologies. However, behind the practical realization and application of these fascinating modern technologies are fundamental principles grounded in the very foundations of physical chemistry, principles that were painstakingly established back in the 19th century, highlighting the enduring legacy of past scientific advancements as they converge with contemporary innovations in analytical techniques [38, 77, 76, 73, 137, 138, 111, 35, 139, 94].

3.3 Detectors

As a mass spectrometer functions primarily as an ion counting device, it necessitates the intricate integration of highly specialized detectors that can effectively convert ions into measurable signals that researchers can analyze and study in further detail. There are several different types of detectors employed in various kinds of mass spectrometers, but among them, the most commonly utilized and widely recognized are Electron Multipliers (EMs) and Faraday Cups (FCs). An Electron Multiplier consists of a carefully arranged series of charged plates, commonly referred to as dynodes, which work together interactively to multiply the input current with each dynode that the ion encounters along its path. When a single ion collides with the first dynode within this multipliers structure, it triggers the rapid release of an electron. This initially emitted electron is immediately propelled toward the second dynode due to its inherent charge. Once the electron strikes the second dynode, the entire process of electron generation and amplification repeats itself and greatly amplifies the signal, resulting in a significant cascade effect that ultimately generates a remarkably robust signal by the time it reaches the final dynode in the sequence. The transit time constant (TC) associated with each unique configuration of these detectors is typically around 1,000. For enhanced clarity when discussing performance metrics in detail, it is therefore more useful to refer to the pertinent concept of effective time to maximum signal for quadrupole mass filters (denoted as ttMSF) and also to use the term time to complete signal for all other types of detectors (designated as tCS). When we analyze the mass spectrometry data further and interpret it with precision, it is determined that at the moment represented by ttMSF plus four times tCS, a truly remarkable 95% of the maximum number of ions has successfully traveled to the detector, thus confirming the overall efficiency and efficacy of the detection process employed in these advanced analytical instruments that are widely applied in research laboratories worldwide [84, 140, 76, 141, 142, 75, 143, 144, 145].

Common detectors that are widely utilized in a diverse array of applications include Channeltrons, which typically display a total count (TC) of around 100. In contrast, Microchannel Plates (MCPs) can reach an impressive TC of approximately 300, showcasing their superior sensitivity and performance in numerous fields. The design and construction of Channeltrons consist of numerous intricate internal channels, which are often ingeniously arranged in configurations that include either 10 or 20 channels to optimize functionality. This specific structural arrangement significantly enhances their operational efficiency by ensuring that ions can interact optimally with the internal surfaces, thus maximizing detection capabilities. When ions make their way into these specialized detectors, they initiate a complex and dynamic process of ionization that results in the effective release of electrons. These released electrons collectively contribute to a series of combined signals, which together create multiple measurable time points, adding to the overall reliability and precision of the measurements. This distinctive characteristic, characterized by responsive and reliable performance, renders Channeltrons a particularly cost-effective and advantageous choice for educational laboratories that are in search of reliable detection technology without incurring exorbitant costs that could potentially hinder their limited financial resources. This is especially critical in educational settings, where budget constraints often lead to tough decisions regarding resource allocation. Therefore, the integration of Channeltrons into laboratory environments not only enhances practical learning experiences but also adheres to financial realities, making them an exceptional choice in the realm of detection technology. [146, 147, 148, 149, 150, 151]On the other hand, MCPs function in a manner that is strikingly similar to that of Electron Multipliers (EMs). Ions that enter the microchannel tubes do so at precise moments when the signal corresponds to a defined left-right (x) and up-down (y) position, allowing for accurate and efficient signal extraction. The resulting signals from MCPs can be plotted with great precision on x/y axes, thus lending itself to thorough and detailed data analysis that is crucial in many scientific investigations. It is particularly important to note that only those ions that make contact with the designated signal area generate detectable signals, which can greatly enhance the quality of data obtained from experiments. When compared to Faraday Cups (FCs), EMs usually exhibit a lower total count when they are operating as a single collector. This phenomenon ultimately leads to improved sensitivity in situations where precision is of utmost importance, thereby making EMs invaluable in high-stakes scenarios. Furthermore, EMs present distinct advantages with their remarkable ability to operate effectively under varying acceleration or voltage conditions, all while also exhibiting lower noise levels in their readings-qualities that scientists highly value in their work. In a noteworthy observation, commercially available Secondary Electron Multipliers (SMPs) can achieve a remarkable total count of up to 106. This outstanding capability underscores their effectiveness and versatility across a broad range of scientific applications and experimental setups. The advancements in these sophisticated detection technologies are continually evolving and improving, making them invaluable resources in scientific research and experimentation. This ongoing progress further drives the exploration into increasingly complex phenomena in various scientific fields, paving the way for groundbreaking discoveries and innovations that can expand our understanding of the natural world. Such advancements highlight the importance of investing in and developing detection technologies that can meet the evolving needs of scientific research and experimentation [146, 147, 148, 152, 149, 150, 151, 153, 154, 155].

It is absolutely crucial to emphasize with great importance that it is not the case that all ions present in a sample are reliably detected in a consistent manner. Various other physical processes must be taken into account when discussing the intricacies of ion detection, such as the multifaceted and complex phenomenon of secondary emissions. This specific phenomenon can lead to the unintentional introduction of additional background noise, which significantly complicates the detection process and substantially reduces the clarity and precision of the results obtained. Generally speaking, the detector often requires active cooling mechanisms to maintain optimal performance levels. This cooling is necessary because it can effectively function much like an accelerating plate throughout the operation. Although comprised of metal, it is vitally important to note that conductivity tends to increase along with rising temperature, which can adversely affect both its sensitivity and overall accuracy in detecting ions. Furthermore, the speed of the detector is also a vital and critical consideration in these intricate systems. This is particularly important due to the inherently scanning nature that is typical of most mass analysers utilized in modern scientific applications. In fact, faster detectors are extremely advantageous as they can provide a pre-selected filter analysis that significantly streamlines the detection process, thereby enhancing the performance. This enhancement ultimately leads to improved overall efficiency in identifying and analyzing the ionic components that are present in the sample being studied and examined in depth [156, 157, 158, 159].

Detector calibration holds immense significance within the field, particularly for flow cytometers (FCs) that may possess various cup sizes or encounter ions arriving at different angles. The calibration process is essential for ensuring precision in measurement. Moreover, the process of routine maintenance is of equal importance, especially if the detectors are undergoing frequent use or are of high value in research and applications. Electronics modules (EMs), known for their inherent fragility, can be relatively easy to break under certain circumstances; however, they can be repaired effectively in a well-equipped workshop with the right tools and expertise. It is crucial that the detectors are properly coordinated with the analyser – for instance, only a spatially-resolved detector can hypothetically be utilized alongside the CMS under ideal conditions, which maximizes the effectiveness of the measurements. Several influential developments in detector technology have emerged throughout the years, including notable improvements in detector sensitivity, enhancements in time resolution, and advancements in parallel detection capabilities that allow for more efficient data gathering. Additionally, several recent breakthroughs in the field are highlighted and discussed in detail, showcasing their practical applications. These advancements not only push the boundaries of current technology to unprecedented heights but also lay down a solid foundation for further innovations, ultimately leading to improved performance, accuracy, and efficiency in various applications [42, 160, 161].

Chapter - 4

Sample Preparation for Mass Spectrometry in Medical Diagnostics

In the specialized and intricate field of mass spectrometry, it is widely recognized and understood that the conclusions drawn from any analytical analyses are frequently only as reliable and valid as the quality of the biological sample upon which these analyses are fundamentally based. The potential for inaccuracy and the inherent unreliability present in the results can often be meticulously traced back to an inadequate preparative effort concerning the specific biological matrix being studied or examined. This comprehensive text covers in detail a wide variety of robust strategies and methodologies designed for the effective treatment of biological samples, all with the end goal of thoroughly informing practitioners about the optimal strategies for preparing blood or tissue samples for an in-depth mass spectrometry analysis. Several methods for isolating, concentrating, and purifying essential analytes are carefully explored, as well as methodologies specifically tailored for the thorough purification and precise preparation of samples. These methodologies have been systematically collated, carefully evaluated, and arranged to delineate their significant impact on enhancing the quality, reproducibility, and reliability of the results obtained through mass spectrometric analyses. Within this context, the procedure of sample treatment is identified as a necessary and critical element that exists within the larger analytical process, with the potential to significantly influence the outcomes achieved in mass spectrometry results. Both inadequate and excessively robust biological sample treatment can lead to potential misinterpretations, the introduction of artefacts, and ultimately, incorrect conclusions regarding general data. After the application of a number of different techniques, it is often observed that problems of sample contamination and the leakage of analytes frequently arise, which subsequently renders the resultant findings and conclusions unreliable and less trustworthy. The high variability associated with the diverse sample preparations that are applied, along with a frequently inadequate or unclear description of these processes, makes the comparison of results a challenging endeavor, or in some cases, even an impossible task. Presented in this comprehensive overview are well-considered recommendations on the acquisition of bio-samples that adhere to the right quality standards for analytical purposes, as well as guidance on the meticulous preparation of representative, uncontaminated, and properly concentrated samples intended for analytical scientific purposes. Furthermore, this current scope of research strives to provide a rational and logical selection of optimal methods for sample preparation processes, with a primary focus on increasing the reproducibility and reliability of the resultant findings obtained from mass spectrometry analyses. This chapter also includes an essential and practical guide on the validity and flexibility of various quality control measures employed in the laboratory, since the quality of any biological sample is often a critical and decisive determinant of the overall research outcomes, regardless of the specific analytical method that is ultimately applied in each case. It must be emphasized and clearly stated that scrupulously selected quality physiological fluids are of crucial importance when it comes to maintaining the integrity of the analyses performed in a mass spectrometric context. For the remaining circumstances described, a diverse array of sample preparation procedures, along with the pertinent results of the corresponding studies, are briefly characterized, analyzed, and discussed. Critical comments and insightful observations are made throughout the text, offering valuable perspective from the viewpoint of a clinical biochemist who is engaged in performing mass spectrometric studies within a clinical setting. For readers of higher academic stature, a thorough understanding of biofluid and tissue sampling techniques may lead to a considerable simplification of the preparative actions required or enable the results generated from the mass spectrometry analyses to be more intelligibly comprehended and appropriately interpreted [162, 73, 57, 36, 163, 164, 71, 165, 42].

Chapter - 5

Quantitative Analysis in Mass Spectrometry

Measurement of analyte concentrations utilizing mass spectrometry is not just a fundamental aspect; it is an absolutely essential component of quantitative analysis prevalent across a multitude of scientific and clinical domains. Within the field of clinical diagnostics, this vital measurement translates into a thorough evaluation of concentrations of various critically important substances, such as proteins, metabolites, drugs, and a vast array of other chemicals located within various patient samples. The accuracy of these quantifications stands at the pinnacle of importance; it is absolutely crucial for making sure that patient diagnostics are not only reliable but also effective. This is particularly relevant in the contemporary setting, given that many medical decisions are directly reliant on the results provided by laboratory tests. The success of any treatment or care being administered to a patient often hinges heavily on the correctness, precision, and promptness of these clinical diagnostics [166, 167, 1, 67, 168, 169, 170, 171, 61]. In sufficiently advanced diagnostic laboratories, mass spectrometry-commonly employed in conjunction with liquid chromatography-has become the go-to method for the comprehensive measurement of a diverse and extensive range of chemicals found in multiple patient samples. Within the realm of medical diagnostics, mass spectrometry has historically been tied to the precise measurement of steroid hormones, immunosuppressants, and various toxic substances that pose significant risks to patient health. As the technology surrounding mass spectrometers has seen advancements, and as costs have undergone notable reductions, these devices have extended their applications far beyond just specialized laboratories into everyday clinical practice. Nowadays, bench-top models of mass spectrometers, which are relatively inexpensive and user-friendly, are frequently utilized across the spectrum of clinical laboratories. These state-of-the-art instruments are often calibrated and adapted to develop and execute targeted assays that are specifically designed for the measurement of small sets of proteins, peptides, or metabolites, functioning seamlessly in combination with liquid chromatography to achieve enhanced efficiency and reliable outcomes. The accurate quantification of any chemical via mass spectrometry when confronted with a complex biological matrix demands rigorous control over numerous measurement conditions. This includes the precise amount of ionizable substance that is carefully introduced into the ion source of the spectrometer, as well as the effective resolution of the ions produced within the spectrometer's source unit. This resolution is absolutely vital in order to prevent potential interferences that could undermine the reliability of readings obtained from the mass analyzer. These two critical conditions form the foundational components of nearly all methodologies that are employed for the calibration of mass spectrometric measurements across laboratories. By implementing a robust calibration protocol using known quantities of analyte, one guarantees that the quantitative measurements derived from these specific conditions and expertly designed instruments are both dependable and accurate, ensuring that results are consistent and repeatable. Moreover, a calibrated mass spectrometer has the distinct advantage of being readily transportable to alternate locations. Measurements taken with a different but equally calibrated spectrometer will yield results that are numerically equivalent, provided the calibration procedures remain strictly adhered to. This capability is particularly beneficial as it simplifies the often complex process of comparing results from distinct studies that are conducted by various laboratories. The practice of utilizing an internal standard for mass spectrometric measurements proves to be highly advantageous, as calibration can be conducted after the fact, relying solely on the signal that is produced within the mass spectrometer during the analysis phase. In this critical context, it is incredibly important to highlight the calibration standards that are routinely employed in mass spectrometry laboratories to meticulously assess instrument performance and reliability. These standards typically consist of carefully prepared mixtures containing known compounds, which are frequently isotopically labeled to enhance precision and reliability. Such commercially prepared mixtures are expertly designed to mimic the classes of chemicals that are commonly analyzed and are routinely used to evaluate operational performance metrics of the mass spectrometer, as well as the robustness and reliability of the measurement method currently in use. Furthermore, the term "metabolome" is particularly significant and deserves special attention when discussing the complete array of low-molecular-weight metabolites that exist within living cells. The metabolome serves as the endpoint of numerous intricate interactions between an organism's genetic inheritance and the myriad environmental influences it encounters. This critical aspect makes the metabolome intricately related to the phenotype of the organism. Moreover, many of the diverse compounds that are classified under this umbrella category are readily accessible through various ex vivo means. This includes obtaining crude extracts from biopsies or a variety of biofluids like blood and urine. Beyond the fact that numerous diseases are correlated with alterations in metabolite concentrations, metabolites tend to be significantly easier to measure compared to other biological markers, making them an appealing target for researchers. This compatibility extends seamlessly to various methods such as solution 1H NMR and multiple mass spectrometric techniques, which enhance the potential for detecting these metabolites accurately. Since the initial documented successes in biomarker discovery that employed a metabolomic approach, there has been an exponential increase in the number of metabolomics studies emerging in the scientific community. This growth underscores the potential of metabolomics as an appealing and viable alternative to traditional genomic and proteomic methodologies in the quest for a deeper understanding of complex biological systems and their underlying mechanisms [172, 173, 174, 166, 175, 176, 111, 2, 177, 35, 37, 178].

5.1 Calibration Methods

Calibration is undeniably an absolutely essential and critically important aspect of measuring devices, encompassing a broad spectrum of complex instruments such as mass spectrometers, all with the primary aim to ensure the utmost accuracy, reliability, and integrity of analytical measurements and results. Mass spectrometry is certainly no exception to this established and significant rule of calibration. Although, in theory, mass spectrometry possesses the remarkable potential to provide highly precise and accurate measurements, this elevated level of precision and accuracy cannot be truly guaranteed or fully realized without the diligent, systematic implementation of proper calibration practices and procedures. Thus, various sophisticated calibration methods, techniques, and mechanisms have been developed, refined, and standardized over time, becoming routine practices adopted and utilized by professionals working diligently within the field. Moreover, it is crucial to note that the inherent bias which often accompanies analytical measurements can be effectively mitigated and significantly reduced through the employment of a thorough and thoroughly well-planned calibration strategy. In addition, the inclusion of pertinent real-world data and comprehensive, illustrative examples will serve to clearly exemplify and highlight the considerable drawbacks, inadequacies, and limitations associated with non-calibrated measurements in analytical practices, ultimately reinforcing the undeniable necessity of rigorous calibration efforts in achieving dependable, accurate, and valid analytical outcomes that can be trusted implicitly [179, 180, 181, 71, 42, 35, 73, 182, 4].

In the specialized field of mass spectrometry, a multitude of calibration strategies is predominantly employed to guarantee both accuracy and precision in various measurements. Among these, the most straightforward and widely recognized method available is the external calibration technique. This approach operates on the fundamental principle that the samples which are intended for analysis are meticulously scrutinized alongside uncomplicated reference samples, which consist of the basic components of the mixture being examined, allowing for a clear comparison. Another widely regarded and effective approach is the use of an internal standard, wherein a select group of known components is carefully maintained at a constant concentration throughout the samples being analyzed. This particular methodology permits the intensities of metabolites present in unknown samples to be evaluated in relation to the intensities of these internal standards, thereby significantly improving the reliability and accuracy of the results obtained. Additionally, there is a specific calibration method that involves the incorporation of matrix samples, which consist of solutions with known concentrations of the substances in question, directly integrated into the analysis process. This form of calibration can enhance accuracy and contribute positively to the overall quality of the measurements derived. However, when we focus on the context of biological specimens, it becomes evident that preparing certified reference materials has proven to be quite impractical due to the inherent challenges it presents; this process can be exceptionally difficult and considerably time-consuming. As a direct result of these complications, it becomes of utmost importance to direct increased attention and focus toward the various possibilities that are associated with the calibration modes previously described. When diligently implementing these outlined procedures in the realm of mass spectrometry, it is vital to give careful consideration to the nature of the sample types that are being utilized. For instance, in cases involving dried liquid samples, solely relying on the external calibration method may inadvertently introduce a bias that can negatively affect the overall accuracy of the test results obtained. Such biases often emerge due to the inherent homogeneity issues associated with solid samples that complicate the analysis process. Therefore, careful and detailed consideration of the chosen calibration method, along with an analysis of the sample characteristics, is absolutely essential for attaining robust and reliable mass spectrometric analyses that can be counted upon for meaningful scientific insights [180, 42, 73, 74, 57, 126, 71, 122, 36].

5.2 Internal standards

Quantitation of analytes through the sophisticated and intricate technique of mass spectrometry is significantly enhanced by the strategic and thoughtful application of meticulously selected internal standards. These internal standards serve as critical and reliable reference points, which are truly indispensable and vital for the precise, accurate, and consistent quantitation of analytes across a remarkably diverse range of samples. Furthermore, these internal standards play an incredibly important role in systematically correcting for variability that can arise from a myriad of factors, encompassing a variety of differences among diverse samples, as well as numerous potential sources of error that may unexpectedly occur during both sample preparation and instrument performance. The deliberate and meticulous inclusion of these internal standards thus provides a dependable, systematic, and highly effective methodology to address, manage, and correct for these variabilities, ensuring consistency and reproducibility throughout the entire analytical process in mass spectrometry. By implementing these internal standards, they serve to assist significantly in controlling the quantification of the analyte by using a second, closely related species that is influenced in exactly the same manner as the analyte itself. This thoughtful approach dramatically enhances the reliability of the analysis and the outcomes derived from the mass spectrometry techniques that are applied in various analytical applications, leading to more robust and confident interpretations of the results obtained through this advanced analytical methodology [183, 184, 185, 186, 187, 188, 189, 190, 42].

The thoughtful selection of an appropriate internal standard hinges on specific criteria that closely align with those of the target analytes. It is imperative to ensure that the chosen internal standards are processed in an identical manner alongside the analytes during each stage of sample preparation and subsequent measurement. This uniform treatment of samples guarantees that the analytes, internal standards, and any additional sources of variability that may arise are affected in a consistent manner throughout the entirety of the analytical process. Consequently, this uniformity is critical to maintaining the integrity of the analytical results. Therefore, even though the signal intensity (SPI) and frequency intensity (FI) may exhibit fluctuations among different samples, these matrix effects will remain constant for both the analyte and its corresponding internal standard. This consistency leads to a reliable compensation for variations that occur during the measurements. As a result, only the concentration changes in the analyte will emerge as a direct function of the specific variable being studied, all while minimizing the influence of interfering factors. This meticulous approach ultimately culminates in improved accuracy and heightened reliability in the quantitative measurements obtained through mass spectrometry, thus reinforcing the importance of internal standards in analytical chemistry [191].

This particular approach brings with it a multitude of advantages, chief among them being its remarkable ability to significantly facilitate the attainment of highly precise and reproducible results. This notable precision is achieved due to the fact that both the internal standard and the analyte are influenced by the same perturbations, thereby allowing for a more controlled and accurate analysis in various conditions. Furthermore, this method necessitates very minimal additional resources in terms of time, cost, and overall effort, making it an ideal choice for researchers and laboratories with limited budgets or resources. Fortunately, there are numerous inexpensive and easy-to-synthesize internal standards that are readily available for use, providing a wide array of options to choose from. The text also provides compelling and illustrative examples that highlight the efficacy of utilizing internal standards for conducting quantitative analyses of organic contaminants in a variety of clinical and environmental samples, demonstrating the versatility and adaptability of this approach. Moreover, it thoroughly addresses several pertinent and critical issues, such as the impact of potential errors that may occur in the concentration of the internal standard, and how these errors can significantly affect the quantitative outcomes when analyzing the concentration of the analyte in question. Additionally, the development of suitable internal standards is discussed in-depth, alongside their critical application to sample types that may present particularly challenging matrix effects, which can often complicate the analytical process and hinder accurate results. Ultimately, the advantages and considerations of this method contribute to a more refined understanding of analytical processes in various scientific fields [192, 193, 194, 195, 196, 197, 198, 199, 200].

Chapter - 6

Qualitative Analysis in Mass Spectrometry

Qualitative analysis in mass spectrometry is fundamentally focused on the critical and nuanced identification of a diverse array of various compounds, alongside the meticulous extraction of essential structural information that pertains to these compounds based on the specific peaks that are inevitably observed in the acquired mass spectra. For many years, the extraction of this structural information was predominantly viewed as the exclusive domain of various advanced spectroscopic techniques, whereas mass spectrometry was traditionally regarded merely as a quantitative tool, primarily centered on providing numerical data that could be effectively utilized in a wide range of analytical evaluations and assessments. It proved to be instrumental in the identification of various compounds, especially in cases where the sample to be analyzed was not readily obtainable in a continuous, consistent, or uniform manner. This particular situation invariably resulted in numerous challenges associated with ensuring a complete, thorough, and accurate analysis of the samples being examined. However, research findings that have emerged over the last few decades have significantly broadened the scope of what can be inferred, deduced, and comprehensively understood from the complex data represented by mass spectra, reshaping the entire landscape of analytical chemistry in profound ways. In contemporary practice, modern sophisticated Spectral Data Handling (SDH) methodologies now empower scientists to recognize, accurately quantitate, and coherently interpret a broad range of functional groups derived from the associated peaks present in the intricate and often convoluted mass spectra. This remarkable advancement has proven to be particularly noteworthy and influential, especially in the indispensable fields of medical diagnostics and in various progressive and innovative "-omics" research areas, such as genomics, proteomics, and metabolomics, which delve into the complex and intricate relationships within biological systems at a molecular level to provide deeper insights. As a result, mass spectrometry has emerged as an invaluable analytical tool, primarily due to its remarkable ability to provide critical structural information about compounds of interest, which is of utmost importance in various applications that span across multiple scientific domains. Consequently, this exceptional capability significantly aids in the effective execution of complex analyses, such as those meticulously observed in the detailed and intricate studies of genomics, thereby enhancing and refining our understanding and application of these advanced analytical techniques in a variety of real-world scenarios across diverse disciplines. This revolutionary development is pivotal in helping researchers draw more meaningful conclusions from their data while simultaneously allowing them to enhance and improve their methodologies in pursuit of new scientific breakthroughs and innovations that can substantially impact our understanding of the world around us [1, 8, 201, 202, 203, 67, 204, 205, 16].

The mass spectrum fundamentally represents the intricate and detailed distribution of ion intensity within a sophisticated mass selective detector. This essential spectrum is intricately linked to the widely varied mass-to-charge ratios of numerous ions that are generated from the sample when it is subjected to meticulous and insightful analysis in a well-calibrated mass spectrometer. There exists an almost limitless and diverse range of possible mass spectra that may be obtained through careful examination and precise analysis; yet, it is crucial to note that for any individual sample, typically only a single mass can be measured at one specific moment during the meticulous and rigorous process of measurement. Consequently, as a result of this inherent limitation in measurement capability, the most reliable and effective methods to confirm a precise molecular identification based on the mass spectrum can be summarized in three critical approaches:

a) Having access to the mass spectrum of a pure sample, which serves as an essential and critical benchmark for comparison purposes,

b) Carefully and systematically comparing the mass spectrum obtained from the sample you are analyzing with that of a known reference compound that has been thoroughly characterized and documented.

c) Employing the mass spectrum to extract detailed structural insights that convincingly validate the identity of the molecule in question.

This systematic method thereby ensures a strong, robust, and thorough identification process. Thus, the overarching objective of qualitative analysis within this intricate and complex realm is to derive precise and accurate measurements from a mass spectrum, which can then be effectively utilized to verify a particular molecular identification with a high degree of confidence, precision, and reliability across multiple analyses. By diligently following these comprehensive guidelines, analysts can significantly enhance the trustworthiness of their molecular identifications and thereby substantially improve the overall outcomes of their analytical endeavors in various scientific applications [73, 206, 207, 208, 209, 45, 210, 211, 36]. In biological systems, the analysis and interpretation of mass spectra can become particularly complex and multifaceted. This complexity arises because numerous different compounds may coexist in the same sample at any given time, making mixtures a common and often challenging occurrence during analysis. Moreover, many of the compounds of interest within these biological contexts may either remain unknown or poorly characterized, thus rendering it impractical to refer to a standard reference mass spectrum for accurate comparison, further complicating the analytical task. A common and effective strategy employed during the analysis of mass spectra involves comparing the acquired spectrum against those stored within a pertinent and well-structured database, which can provide crucial context and reference points. With the implementation of suitable analytical methods, the probability of encountering a random match, which results in a false positive identification, can be significantly minimized. As a result, the practice of utilizing spectrum searching against established databases has become one of the most frequently adopted and reliable analytical strategies in the ever-evolving realm of organic mass spectrometry (OMS). While qualitative analysis can sometimes be prone to misinterpretations and errors in the interpretation process, it is crucial for analysts to exercise caution when interpreting the mass spectrum of any particular compound, as even small oversights can lead to significant consequences in identification accuracy. The incorrect assignment of peaks is a common pitfall in this field, as it can happen quite easily and is often overlooked, especially during high-paced analyses. Nevertheless, qualitative analysis possesses a vast array of potential advantages that can greatly enhance research outcomes. There is a significant amount that can be achieved when it comes to leveraging the rich source of information contained within any given mass spectrum. Part of this enhancement in analysis can be attributed to the availability of comprehensive databases along with specialized tools that assist in facilitating and automating critical analytical tasks. Such tasks encompass the preparation of mass spectrometry (MS) data for analysis, the systematic searching of spectra against extensive databases, as well as the prediction and interpretation of fragment ions derived from the mass spectrometric evaluation. These advancements, driven by modern technology, have the potential to considerably augment the reliability and accuracy of qualitative assessments in mass spectrometry, paving the way for more insightful discoveries [212, 213, 214, 215, 216, 217, 218, 219, 220, 207, 221].

6.1 Database Searching

Database searching represents an essential and fundamental technique in the expansive realm of qualitative mass spectrometry, acting as a cornerstone that promotes and enhances the effective identification of numerous compounds or, at the very least, assists in narrowing down the vast options to appropriate classes of compounds. Over the years, a multitude of specialized databases have been meticulously established explicitly to support the identification and facilitate the in-depth analysis of small molecules. These comprehensive, well-curated databases offer reference spectra of pure compounds, ideally measured under the identical conditions that are utilized for the analysis of relevant biological samples, ensuring an essential consistency and reliability across various research endeavors. Furthermore, software tools are fundamental in achieving an efficient and accurate search through the thousands of entries housed within a compound database. Such tools empower researchers to effectively identify the specific compounds present in a given sample while meticulously confirming diagnostic signals that have been detected during analysis. With the assistance of sophisticated and advanced database-search algorithms, analytical workflows within laboratories are markedly enhanced: conventional parameters must be judiciously and carefully selected according to a well-defined and structured set of rules. This process not only streamlines the entire workflow but also ensures the pre-selection of a small group of candidate compounds that may warrant more thorough investigation. As a direct result of these remarkable advancements in technology, only a limited number of compounds require further verification during the subsequent and often more time-consuming step of qualitative analysis. This crucial refinement contributes to an overall increase in both efficiency and productivity in the laboratory setting, effectively enabling scientists to focus their valuable efforts on the most promising candidates for further study. The implementation of these cutting-edge search techniques plays a pivotal role in facilitating a more precise identification process, ultimately leading to improved outcomes in research and development involving small molecules across the diverse spectrum of various scientific fields. The remarkable impact of these endeavors can be observed in multiple applications, ranging widely from drug discovery to environmental monitoring, where reliable identification proves to be critical in advancing knowledge and driving innovation [212, 222, 2, 223, 32, 224, 225, 226].

However, the crucial parameters for the identification process-specifically, the entry of the mass spectrum as well as the search algorithm used in the process itself-can hardly ever be reliably predicted with confidence. This inherent unpredictability fundamentally makes database searches susceptible to various inaccuracies and potential errors that could undermine the integrity of the results. Given the impressive and significant potential associated with the ever-evolving mass-spectral databases, it becomes increasingly crucial to execute and interpret searches in a highly disciplined and methodical manner. This meticulous approach involves undertaking a thorough and careful validation of all the results obtained from such searches to ensure their overall reliability and correctness. Ultimately, the quality of the mass spectrometry (MS) data, particularly focusing with great attention on the accuracy of the collisionally induced dissociation (CID) spectra, remains the cornerstone of a successful and unambiguous assignment process of small molecules in analytical chemistry. This principle holds true, even with the clever and proficient utilization of various databases that have been developed. Several compelling examples illustrate how databases are currently being effectively leveraged in the rapidly growing and dynamic field of metabolomics. Additionally, the new technologies and innovative search strategies that are being actively pursued and developed by various vendors are also presented for discussion. The strategic integration of standardized databases designed specifically for MS data is expected to not only enhance the overall acceptance and credibility of mass spectrometry across different scientific and research fields but also facilitate a much more seamless exchange of valuable data between users. This encourages collaboration and advancement within this crucial area of scientific research and development [227, 228, 229, 230, 231, 232, 233, 18, 16, 234, 235].

6.2 Structural elucidation

6.2.1 Fundamentals

In addition to the optimal utilization of accurate mass, a multitude of other analytical techniques are employed for a comprehensive structural elucidation of the complex molecular structures associated with various types of chemical compounds. As the continual development and significant advancement of modern mass spectrometers persist, mass spectrometry has evolved into a vital instrument that is indispensable for both routine and detailed structural characterization of diverse, complex molecules. Within the intricate mass spectra of various molecules, there exists a substantial amount of important and unique information, which includes accurate molecular mass, detailed fragmentation patterns, distinct isotopic distributions, and the specific sites of isotopic labeling that can be observed. These critical data points are fundamental for greatly enhancing the understanding and interpretation of molecular structures. For instance, the presence and significance of aromaticity within a molecule can be confirmed through the use of the well-established A-rule; additionally, the presence of nitrogen (N) in a compound's structure can be inferred from a larger molecule losing a fragment that corresponds to 28 Da in mass; furthermore, the distance between two heavy atoms can be estimated by closely examining the difference in the movement of a deuterium atom that occurs during the complex process of drug metabolism. The thorough elucidation of molecular structures is indeed an essential step in the elaborate analysis and handling of complex samples found within the fields of clinical diagnostics research or the important realm of drug development. This elucidation can be accomplished through a variety of strategies, such as the careful interpretation of significant peaks found within the spectra, conducting detailed spectral matching between experimental data and reference data, as well as comparing proposed structures or substructures against entries found in relevant comprehensive databases, or deducing the structure of the compound by following a specific set of well-established, systematic rules. More recent and sophisticated strategies have also been developed specifically for the interpretation of mass spectral signatures, which continue to allow researchers to provide additional vital structural endpoints, such as the correct identification of potential metabolites, the accurate delineation of characteristic neutral losses associated with structures, or the recognition of stable fragment ions that play a role in the analysis. Common molecular structures that typically appear in a vast variety of biological molecules are generally analyzed and understood through comparative examples, encompassing comparisons like Phosphate versus Sulfate or Drug+H versus Drug+Na versus Drug+K, which highlight the different behaviors and interactions of such molecules. It is crucial, furthermore, to also discuss any potential artifacts that may occur or be introduced throughout the various analytical processes used. As illustrated in the detailed examples presented below, structural elucidation that is derived from the resulting mass spectra is an absolutely critical step within the intricate realms of untargeted or non-targeted metabolomics or proteomics studies and plays a significant role in enhancing the understanding of the underlying chemistry of the mass spectrometry techniques that are involved. This encompasses comprehensive discussions that revolve around effective strategies for the structural analysis of typical mass spectral signatures and includes practical examples drawn from real-world applications. In the field of mass spectrometry, detailed molecular structure information is gleaned through a thorough analysis of its spectra. This intricate aspect becomes considerably more challenging when addressing high-throughput proteomics and metabolomics technologies, which generate an extensive array of analytes exhibiting highly complex structures, often presenting in the form of intricate mixtures that require careful separation and identification. This expanding section will detail an in-depth discussion that focuses on the implications of fragmentation and the deeper meaning of mass spectral signatures relevant for the analysis of various types of molecules. It will include a comprehensive description of strategies for conducting structural analyses that encompass interpreting mass spectral signatures, thoroughly deciphering peptide sequences (including the identification of various modifications), translating the significant findings to relevant molecular formulas, and ultimately leading to the inference of the overall molecular structure. Both specific examples and general principles regarding how mass spectra can effectively be interpreted and used for structural elucidation will be examined in depth, providing a comprehensive overview. Furthermore, there will also be a discussion involving active collaborations with mass spectroscopists regarding the diverse strategies deployed in structural elucidation, along with insightful perspectives into the future directions and potential advancements of the field of mass spectrometry. Alongside these insightful discussions of molecular structure, it is pertinent to continuously address the validation of any proposed structures, a task that proves to be quite difficult in the complex realm of mass spectrometry, yet it becomes significantly easier when it is feasible to conduct advanced mass spectrometric analyses on isotopically labeled samples. For compounds that are typically found with a certain substituent that is located at a designated single position, extra care is taken to deeply consider compounds that, by definition, are classified as constitutional isomers. With the remarkable onset of high-resolution mass spectrometers, it has become increasingly possible to accurately identify a wide range of molecules based on their respective molecular formulas as well as the exact mass of the ion species that are observed in detailed analyses. Recently, advancements and innovations in high-resolution mass spectrometers have enabled researchers to identify adducted molecules and those that are isotopically labeled with remarkable precision, contributing greatly to the field's continuing advancement and understanding [212, 74, 122, 124, 236, 237, 125, 34, 71, 209].

It is absolutely crucial to clearly state any potential limitations that may arise in the comprehension of a spectrum, as such limitations can significantly hinder accurate interpretation and analysis in various scientific contexts. Furthermore, it is fundamentally important not only to explain effective strategies on how to avoid these pitfalls but also to provide practical guidance on how to better interpret the problematic spectra that may lead to confusion or misinterpretation among researchers, analysts, and those engaged in scientific exploration and inquiry. These interpretations are often susceptible to errors, and thus a comprehensive understanding becomes vital for effective scientific communication. Additionally, the incorporation of other analytical techniques, such as Nuclear Magnetic Resonance (NMR) or Fourier Transform Infrared Spectroscopy (FT-IR), is seriously taken into consideration in the holistic and comprehensive discussion of mass spectral elucidation. This consideration aims to enhance the accuracy and reliability of results obtained from mass spectrometry, ensuring that researchers can achieve greater precision in their findings and conclusions drawn from such data. A dedicated sidebar specifically mentions various software packages that are currently available, or are anticipated to become readily available in the near future, which have been expressly designed to calculate or assist in the structural elucidation process from mass spectra. These innovative tools and software are invaluable resources for researchers who are aiming to enhance their understanding and interpretation of complex mass spectral data, thereby facilitating more reliable and accurate outcomes in their analytical endeavors. In this context, it becomes clear that adopting a multi-faceted and integrative approach, which combines various analytical methods and modern computational tools, is essential for effectively overcoming the considerable challenges posed by complex spectral data. Thus, allowing for a greater understanding and improved analytical capabilities in the ever-evolving field of scientific exploration and research, ultimately advancing our knowledge and capabilities in analytical chemistry and related disciplines [238, 126, 73, 239, 36, 74, 180, 240].

Chapter - 7

Applications of Mass Spectrometry in Clinical Chemistry

Mass spectrometry stands out as an extraordinarily powerful and remarkably versatile analytical technique that is decidedly essential for the thorough analysis of a vast and diverse array of molecules. This extensive range includes, but is by no means limited to, proteins, peptides, and various metabolites. Over the last few decades, the significance and application of mass spectrometry have escalated dramatically within the critical realm of clinical chemistry laboratories. A clinical chemistry laboratory urgently requires reliable and robust analytical methods that enable the precise measurement of various biomolecules. These biomolecules encompass proteins, metabolites, and lipids, all of which can be found in complex biological matrices that often present significant challenges for achieving analytical accuracy and precision. Within this intricate and dynamic setting, mass spectrometry holds a prominent and vital position among the plethora of measurement methods that are routinely implemented in standard clinical chemistry practices. In fact, it is particularly noteworthy that at present, approximately half of all analyses conducted in clinical chemistry laboratories depend on the sophisticated techniques associated with mass spectrometry. This invaluable analytical method possesses the remarkable capability not only to accurately characterize individual compounds with exceptional specificity and detail but also to effectively screen for multiple distinct compounds in a simultaneous fashion. The versatility inherent in mass spectrometry makes it widely utilized across a diverse range of applications. These applications include the rapidly advancing fields of proteomics, metabolomics, and lipidomics. All of these areas have attracted substantial attention and growing importance in recent years, both within the context of scientific research and in the domain of clinical diagnostics, as they contribute critically to advancing our understanding of complex biological systems [241, 168, 61, 165, 242, 243, 244, 245, 246, 247, 2].

Proteomics is an intricate and specialized field that delves deeply into the comprehensive, complex, and detailed analysis of the myriad proteins expressed by a given genome. Each of these proteins can be significantly influenced or altered by a wide array of environmental, physiological, or behavioral factors that an organism encounters throughout its entire lifecycle. This particular area of study is not only highly valuable but incredibly important, as it facilitates the discovery of crucial proteins that are essential for effectively predicting and understanding disease aetiology, as well as the progression and development of various diseases, and the responses to an extensive range of therapeutic interventions that might be applied. Within the confines of clinical chemistry laboratories, the application of advanced mass spectrometry techniques has grown tremendously significant and pivotal, particularly when it comes to the accurate determination of biological markers that correlate directly with various diseases. These biological markers serve as critical indicators, playing a fundamental role in monitoring the recurrence of tumors following surgical excision and effectively tracking the numerous physiological changes that may occur in individuals post-therapy, thus providing vital information that can guide further treatment decisions and improve patient outcomes. This makes the research and results within the realm of proteomics invaluable for the future of personalized medicine and targeted therapeutic strategies [42, 128, 45, 1, 25, 248, 3, 249, 10].

In contrast, metabolomics embodies a comprehensive systems biology approach that meticulously focuses intently on the in-depth analysis of the levels, concentrations, and fluxes of small molecules that are known as metabolites. Much like the field of proteomics, mass spectrometry stands out as a prominent and widely utilized technology within the expansive and ever-evolving domain of metabolomics. Notably, in various scientific situations where a myriad of probes are carefully and strategically labeled with stable isotopes, mass spectrometry can be effectively integrated with a diverse and extensive range of other detection methods. This integration leads to the development of robust, comprehensive, and reliable metabolomics assays that can yield incredibly insightful data. However, it is worth mentioning that NMR-based metabolomics often grapples with substantial challenges and a multitude of limitations concerning its lower sensitivity. This limitation is particularly stark when juxtaposed with the remarkable capabilities and advantages that are offered by mass spectrometry technology. Consequently, mass spectrometry tends to be favored and preferred by analysts as well as researchers who are earnestly pursuing these progressive and innovative methodologies in their scientific work, greatly appreciating the efficiency and effectiveness it brings to their studies and the pursuit of knowledge [8, 34, 250, 251, 37, 57, 252, 135, 54].

Additionally, a closely related field is lipidomics, which involves the expansive, systematic, and comprehensive analysis of all lipids contained within a biological system. To this end, researchers have employed advanced gas chromatography techniques in conjunction with time-of-flight mass spectrometry (GC-tof-MS) to meticulously investigate the diverse lipids found in blood serum samples. Through their thorough and in-depth studies, they have successfully identified specific lipids that appear in elevated concentrations among individuals who are at a heightened risk of developing diabetes. Furthermore, a rapid diagnosis kit has been designed and proposed; this innovative tool utilizes a quick and efficient mass spectrometric method to accurately measure the contents of several critical lipids, thus significantly enhancing the efficiency and accuracy of diagnoses related to metabolic disorders. This notable advancement in technology represents a monumental leap forward in how metabolic disorders are diagnosed and managed within the clinical realm, ultimately improving patient care and outcomes [61, 42, 128, 253, 10, 254, 1, 3, 255, 256].

7.1 Proteomics

Mass spectrometry has experienced truly remarkable advancements over recent years and has made significant inroads into a multitude of crucial fields related to both research and diagnostics. These vital fields encompass essential areas that include, but are certainly not limited to, the essential disciplines of genomics, proteomics, lipidomics, drug analysis, and broad-ranging assessments that are fundamentally related to food safety and environmental integrity. Furthermore, mass spectrometry plays an absolutely essential role in the important realm of medical diagnostics. This extraordinary progress is a direct product of both the critical requirements and essential needs demanded by these various domains, as well as the powerful and versatile analytical capabilities that state-of-the-art mass spectrometers provide for researchers and laboratories alike. Specifically, mass spectrometry, along with its closely related and critically important relatives, most notably the expansive fields of proteomics and clinical proteomics, has emerged as an invaluable and vital tool in the far-reaching realms of biomedical research and clinical laboratories. The sophisticated methodologies associated with mass spectrometry provide unprecedented and highly advanced capabilities that allow for incredibly detailed and in-depth analysis of complex biological samples. These advanced and innovative methodologies possess significant potential to unmistakably revolutionize various aspects of medical practice by offering transformative insights and solutions that were previously considered unattainable within the medical community. The integration of mass spectrometry into diagnostic and therapeutic processes not only significantly enhances the accuracy of vital measurements but also leads to improved patient outcomes and opens expansive new avenues for scientific investigation and discovery. With the relentless evolution of these cutting-edge technologies, the future of mass spectrometry holds the exceptional promise of even greater breakthroughs and extensive applications across a variety of sectors, ensuring its enduring position as a cornerstone of modern analytical science and practice. [16, 257, 258, 171, 259, 260, 261, 262]This chapter aims to provide a thorough and insightful introduction to the fundamental principles as well as the essential terminology that are intimately associated with the complex, intricate, and multifaceted field of mass spectrometry. Furthermore, it delves into an extensive variety of significant aspects and influential applications of mass spectrometry that are currently prevalent within the life sciences sector today. It places a strong emphasis on its current and potential applications within the ever-evolving, dynamic realm of infectious disease studies. Recent technological advancements in the field have now made it possible to meticulously, accurately, and efficiently analyze crucial genetic materials such as DNA, RNA, and proteins through the use of highly sensitive and powerful analytical techniques. Among these leading techniques, mass spectrometry (MS) stands out prominently, as it offers a remarkably detailed and quantitative mapping of individual genomes, transcriptomes, and proteomes at an unprecedented level of precision. This precise quantitative mapping is instrumental not only in revolutionizing the expansive field of biomedicine but also in greatly enhancing our understanding of fundamental biological processes and their intricate complexities. By elucidating and unraveling these complexities, mass spectrometry plays a pivotal and essential role in advancing medical research, as well as in significantly improving diagnostic methods. This, in turn, fosters the development of innovative therapeutic strategies in the ongoing fight against various diseases that challenge public health and well-being across the globe [1, 24, 2, 263, 28, 59, 264, 60].

Proteomics, in essence, is characterized as the large-scale study of proteins, with a particular focus on understanding their intricate structures and multifaceted functions, as well as the complex ways in which they are synthesized, degraded, and modified within different cells and tissues. Essentially, proteomics involves a comprehensive and expansive analysis of proteins on a large scale, which includes the structural and functional examination of the entire repertoire of proteins that are expressed within a specific cell, tissue, organ, organism, or biological fluid, all particularly under designated physiological or pathophysiological conditions. It plays a critical role in determining the unique structural configurations of all proteins present within a given genome and in identifying all pertinent protein-protein interactions that occur within a specific organism or cellular system. The driving force behind the relentless pursuit of enhanced diagnostics, coupled with an improved comprehension of the complex molecular mechanisms that underlie various diseases, is elaborated upon in this text, along with a thorough review of the implications this has on the future landscape and promising developments within personalized medicine. This chapter further outlines various innovative methodologies and approaches employed for the intricate analysis of proteins through the lens of mass spectrometry, providing a comprehensive overview that encompasses not only sample processing techniques but also various methodological approaches for effective protein separation, the robust methods for protein fragmentation, efficient mass spectrometric detection protocols, and the sophisticated identification processes for proteins. Despite the rapidly growing interest and significant engagement in the dynamically evolving field of proteomics, it is noteworthy that the number of therapeutic drugs based on proteins remains inconceivably smaller in comparison to the vast number that have been developed from small molecule therapeutics. The chapter delves deeply into the utilization of proteomics in drug development and highlights several successful case studies alongside practical examples that have been meticulously documented within the scientific literature up to this point. Broad-spectrum proteomics studies have only been able to identify a noticeably smaller subset of individual proteins or interactions between proteins, signaling that much remains to be elucidated and thoroughly understood within this continuously evolving field. Several of the most pressing challenges currently confronting researchers and practitioners within this realm are comprehensively discussed in this overview, particularly those challenges that relate to the inherent complexity of protein mixtures, the diverse abundance of various proteins, the numerous post-translational modifications that proteins can undergo, alongside the variability observed in samples and their general representation. The text also offers thoughtful suggestions regarding how these pressing issues might be effectively addressed and conquered in future research endeavors to ensure steady progress in the field. In conclusion, emerging trends within the proteomics landscape are thoroughly explored, providing insightful observations about the development of new, exceedingly more sensitive, and selective mass analyzers, innovative labeling strategies, and cutting-edge instruments that are designed specifically for the accurate measurement of global protein turnover within biological systems. Additionally, the text considers high-definition mass spectrometric analysis of post-translational modifications, along with a variety of applications within chemical biology and systems biology, and mentions advancements in the creation of new libraries of antibodies that target specific proteins. Moreover, it emphasizes the ongoing evolution of diagnostics and proteomic analyses that are closely associated with disease predispositions and their overall impact on enhancing patient care [241, 265, 18, 1, 258, 266, 267, 175, 268, 269, 270, 271].

7.2 Metabolomics

Mass spectrometry is increasingly gaining significant interest in the rapidly advancing field of metabolomics, which is defined as the systematic study and analysis of small molecules, commonly referred to as metabolites. These metabolites are present within the cells, biofluids, tissues, and various organs of living organisms, serving crucial roles in numerous biological processes. The technologies associated with mass spectrometry indeed play critical and essential roles in the overall domain of metabolomics, primarily due to their inherent detection sensitivity and specificity. This remarkable sensitivity enables researchers to detect, identify, and quantify metabolites with a high degree of accuracy, finesse, and precision that is vital for understanding biological mechanisms at a profound level. Metabolomics-related mass spectrometry methodologies and applications are thoroughly outlined as follows: Firstly, how mass spectrometry can significantly assist researchers and scientists in enhancing our understanding of the incredibly complex metabolic processes occurring within a biological system and how these processes influence overall health. Secondly, how to effectively apply the advanced and cutting-edge mass spectrometry technologies to profile and study various metabolic diseases that affect countless individuals worldwide, providing crucial insights into disease progression and potential therapeutic targets. Lastly, how the discovery and validation of critical biomarkers for metabolic diseases, such as diabetes, obesity, and cancer, can be effectively achieved through the utilization of sophisticated mass spectrometry technologies, paving the way for innovative diagnostic tools and targeted therapies in modern medicine. This advancement holds the potential to revolutionize metabolic disease management and enhance overall health outcomes significantly, transforming the landscape of personalized medicine and improving the quality of life for many individuals struggling with these conditions [272, 8, 250, 251, 273, 274, 275, 276, 71, 277].

Metabolomics is an emerging and rapidly evolving branch of the extensive field of “omics” sciences that has gained significant attention and focus in the ever-changing landscape of the post-genomic era. Its importance is continuously increasing and growing, particularly for the purposes of conducting comprehensive and holistic studies that delve into complex living systems and their intricate interactions. The primary aim of metabolomics is to broaden the understanding of various phenotypes specifically at the level of the metabolome. The metabolome refers to the entire and complete set of numerous small molecules that exist within a biological system, which can encompass a wide variety of substances found within conducting mediums, cellular extracts, and an array of different tissues throughout an organism. Recent studies have made noteworthy and promising attempts to utilize advanced and cutting-edge mass spectrometry technologies in order to analyze rare and/or particularly complex samples that are extracted directly from solid tissues. The analytical methodologies that are built upon the principles of mass spectrometry have been meticulously and carefully designed for the thorough and comprehensive analysis of metabolites in those samples. These methodologies include several hyphenated technologies such as Gas Chromatography-Mass Spectrometry (GC/MS), Liquid Chromatography-Mass Spectrometry (LC/MS), and Capillary Electrophoresis-Mass Spectrometry (CE/MS), in addition to the extensive utilization and application of mass spectral databases that significantly enhance both the metabolite identification process and the quantification processes as well. Mass spectrometry, in particular, plays a pivotal and crucial role in propelling research initiatives and driving new discoveries within the expansive field of metabolomics. Ongoing advancements and innovative applications related to clinical practices are set to revolutionize the field of analyte profiling, potentially transforming it into a routine and standardized analytical discipline commonly used in laboratories. Nevertheless, it is important to acknowledge that many challenges still persist within the metabolomics domain. These challenges comprise factors such as sample complexity, analyte stability, and the throughput of various essential processes. Continued and sustained advancement in both mass spectrometry technologies as well as robust data analysis methodologies is critically required in order to tackle the pressing issues or, at the very least, diminish some of the obstacles that currently limit the broader application and successful integration of metabolomics into scientific research and clinical settings [8, 71, 251, 203, 278, 279, 250, 280, 276].

7.3 Lipidomics

Mass spectrometry (MS) has long been utilized for the determination of lipids, which constitute an intricately complex group of molecules that includes triglycerides, phospholipids, sterols, and a variety of other organic macromolecules that are typically found within various biological samples. These lipids play an incredibly important and multifaceted role in both health and disease due to their diverse range of biological functions. The unique and advantageous property of mass spectrometry lies in its exceptional suitability for accurately determining a wide array of compounds with a broad range of polarity, which effectively allows researchers to analyze differing types of lipids in a highly efficient manner. The extensive and remarkable structural diversity inherent in lipid molecules renders them considerably more complex compared to most metabolites or protein compounds typically encountered in biological research. When one examines the myriad possibilities for variations occurring along both the fatty acid chains and the head groups, it swiftly becomes apparent that it is indeed possible to assign a distinct identity to only a limited fraction of the various peaks observed in the typical recorded ESI mass spectrum of human plasma. This complexity presents a significant challenge in the comprehensive and thorough analysis of lipids, making it a pursuit that necessitates both time and careful attention to detail. Numerous publications have meticulously reviewed the crucial role of mass spectrometry in the preparation and analysis of lipid processes across multiple domains, including metabolic, cardiovascular, nutritional, genetic, and environmental diseases, further highlighting the breadth of its application. These reviews are well documented in the scientific literature, and this work aims to present a thorough and insightful review of the current methodologies being employed in the field, while also emphasizing the urgent and pressing need for improved clinical applications that can standardize the overall lipid analysis methods utilized in modern practice. Lipids play multiple critical and essential roles in various cellular processes; furthermore, the disorders related to them have been increasingly implicated in a wide range of metabolic and cardiovascular diseases, most notably including diabetes, hypertension, and atherosclerosis. Mass spectrometry not only enables detailed and intricate profiling of representative lipid species but also allows for a comprehensive analysis across different classes of lipids-something that is not easily achieved through other methodological approaches currently available. Within this rapidly evolving field, various methodologies tailored for both targeted and untargeted lipid analysis are presented alongside valuable insights into the practical applications of this analysis within clinical diagnostics. The processes through which lipid analysis can significantly aid in enhancing our understanding of complex cellular processes and unraveling the underlying mechanisms associated with various diseases are considered to be quite general and highly applicable across multiple research domains. Throughout the review, specific results generated using the QTRAP 3200 LC-MS/MS instrument are highlighted, as well as findings from the commercially available Aushon Ciraplex assay for lipid analysis, both emphasizing their relevance and effectiveness in lipid research and robust clinical diagnostics

[281, 282, 283, 284, 285, 286, 201, 56, 287, 54, 7].

Chapter - 8

Mass Spectrometry Imaging in Medical Diagnostics

Introduction

The innovative and cutting-edge application of mass spectrometry (MS) imaging within the dynamic and rapidly growing field of medical diagnostics has emerged as a truly transformative and highly promising area of research and practice, revolutionizing the way we approach and understand disease detection and treatment strategies. This article offers an extensive and in-depth review of the fundamental principles that underpin mass spectrometry, ensuring that readers thoroughly grasp not only the core concepts but also the remarkable intricacies and nuances of this groundbreaking technology. It explores the numerous medical applications, which are not only emerging but also significantly expanding in their importance and clinical relevance, thereby highlighting the multifaceted and diverse nature of this invaluable tool in contemporary medicine and diagnostics. The continual advancements in technology and methodology related to mass spectrometry further contribute to its ability to provide insightful and detailed analyses, ultimately enhancing patient outcomes through improved diagnostic capabilities. By delving deeper into this field, we can better understand the potential of mass spectrometry as it continues to evolve, presenting exciting possibilities for future research and clinical application, particularly across various specialties like oncology, neurology, and infectious diseases. Through ongoing innovation and exploration within the realm of mass spectrometry, researchers and healthcare professionals are equipped to push the boundaries of medical science, paving the way for enhanced diagnostic precision and personalized treatment approaches tailored to individual patient needs. In particular, it presents a detailed focus on biological MS applications that hold immense relevance and particular significance within the clinical setting. These applications illustrate, in comprehensive detail, how diverse methodologies have been effectively integrated into current practices or are anticipated to be incorporated in the near future to serve essential purposes of disease diagnostics. Furthermore, the article delves deeply into the vital and important applications of mass spectrometry within the critical area of pharmacokinetics. This specific domain pertains to the meticulous and in-depth analysis of drugs and their metabolites across various biological samples and different matrices, shedding light on how monitoring drug levels can lead to improved patient care. This comprehensive discussion emphasizes the critical importance of quantitative imaging approaches, highlighting the remarkable advancements in direct analysis technology that continually enhance and broaden the capabilities as well as the potential of mass spectrometry in the ever-evolving field of medical diagnostics. Through this extensive exploration, it becomes increasingly clear how central mass spectrometry has become in the ongoing efforts to improve diagnostic capabilities, streamline processes, and ultimately provide better healthcare outcomes. As the integration of mass spectrometry continues to evolve, its role in shaping the future of medical diagnostics becomes ever more significant, calling for further research and development to harness its full capabilities.

Mass spectrometry imaging (MSI) is an innovative and rapidly progressing branch within the expansive field of mass spectrometry that facilitates the spatially resolved and detailed analysis of complex biological tissues. Since its introduction in the 1990s, MSI has undergone significant growth and evolution, leading to the development of an array of advanced techniques that excel in the direct analysis of biological tissues as well as intricate tissue microarrays. Presently, one of the most widely adopted and utilized approaches to MSI is matrix-assisted laser desorption/ionization, which provides scientists and researchers with the remarkable capability to obtain detailed insights into the intricate distribution patterns of various molecular species present within diverse tissue types. These molecular species encompass a broad range, including small molecules like pharmacological drugs, their metabolites, as well as endogenous metabolites or lipids, peptides, and proteins. This rich and multifaceted molecular information can be effectively complemented by integrating other sophisticated imaging techniques, such as histological evaluations, elemental analyses, or advanced immunohistochemistry methodologies. The integration of these diverse methodologies has significantly enhanced the ability to spatially localize essential biomolecules or metabolites that are relevant to a myriad of pathologies, health conditions, and disease states. This remarkable capability has consequently fostered considerable growth in both therapeutics and diagnostics research, paving the way for novel treatment strategies and improved diagnostic methods. However, despite its numerous advantages, the application of MSI can also present several significant challenges that researchers must navigate. For instance, it is often characterized by relatively low throughput, which can hinder the overall efficiency of analyses and slow down research endeavors. Furthermore, the costs involved in employing MSI can be prohibitively high, particularly when considering the cutting-edge specialized equipment and techniques required. Additionally, the interpretation of the generated data can often be complicated and may necessitate extensive expertise and training. Moreover, in certain molecular channels, the resolution of the data remains insufficient, which can limit the overall reliability and practical utility of the results obtained. These persistent challenges, along with the many ongoing innovations and implementations in the field, currently serve to limit the broader adoption and use of this advanced technique in clinical practice settings, restricting its full potential in enhancing patient outcomes [37, 288, 71, 29, 289, 109, 290, 291].

Chapter - 9

Challenges and Limitations of Mass Spectrometry in Medical Diagnostics

Despite the notable recent advancements and the increasing availability of a wide range of mass spectrometry (MS) platforms, there continue to exist significant challenges and limitations that arise when attempting to effectively deploy mass spectrometry in intricate medical diagnostic workflows. One of the foremost issues encountered is the high sample complexity involved, as the matrix of various biological samples can include a diverse array of fluids such as plasma, sputum, lymph, bile, or urine. Each of these biological fluids harbors a multitude of similar substances, which serves to complicate the overall analysis process considerably. Furthermore, matrix effects embody a common problem, as the presence of various matrix components can not only alter the ionization of the analyte directly, but can also affect the transmission of ions to the mass spectrometer and detectors, often due to contamination that can occur within the sample cone. In addition to these complications, most inorganic salts generate an abundance of ions during the process of mass spectrometric analysis, which heavily weighs down and impacts the overall performance and efficiency of mass spectrometric systems and methodologies. Specifically, the mass spectrometric analysis of complex samples such as whole blood and plasma frequently leads to significant ion suppression, making it particularly challenging to obtain reliable and reproducible readings. The ultimate concentration and the likelihood of efficient ionization are determined by the presence of a vast and immeasurable number of molecules with differing chemical natures and properties that are involved in the intricate analysis process. Given the critical need for proper system operation and the highly specialized skills required for effective data evaluation and interpretation, one of the main barriers hindering the broader utilization of MS technologies in the healthcare sector may well stem from the technical and staff training aspects that are necessary for operating and understanding such advanced analytical platforms. Furthermore, adhering to prevailing regulatory requirements and maintaining a standardization within quality control (QC) processes are essential prerequisites for the successful commercialization of medicinal products, particularly when considering the stringent European CE marking regulations and their implications for securing availability for reimbursement within healthcare systems [21, 292, 96, 293, 294, 295, 296, 297, 298].

Looking ahead towards the future of diagnostic methodologies, the anticipated integration of innovative technologies such as organs-on-chips, advanced liquid biopsy technologies, and cutting-edge bioinformatics tools-including the promising applications of artificial intelligence in diagnostics-might be considered a vital direction and avenue to leverage the current toolsets available within the field. Nonetheless, due to the substantial economic limitations faced by healthcare institutions, which include the significant costs associated with the acquisition of sophisticated diagnostic devices and the related expenses of employing highly trained and qualified staff, the application of these technically advanced medical technologies is likely to be primarily confined to specialized centers that possess the capability to provide such quality services using appropriately trained personnel dedicated to these tasks. The establishment of extensive national clinical centers that are well-equipped with comprehensive social support from various associations of companies could pave the way for substantial improvements in accessing advanced medical technologies. Additionally, broad branches specializing in fields closely related to bio-nanotechnologies-while not exclusively limited to medical devices manufacturers-could act as vital suppliers of such essential equipment for research facilities across various disciplines. These same branches could also serve to supply regional clinical centers with the necessary support from recognized entities like the CEA, while being actively involved in translating and implementing promising research findings into practical applications that can be utilized in patient care. This collaborative and synergistic effort could significantly bolster the wider deployment of mass spectrometry and other advanced diagnostic technologies in clinical settings and practice, ultimately enhancing the overall quality of care that is available to patients in need of these crucial medical services. [23, 74, 57, 299, 36, 126, 300, 301, 109, 302, 237]

Chapter - 10

Future Perspectives and Emerging Trends in Mass Spectrometry for Medical Diagnostics

Mass spectrometry, particularly when it is integrated with its associated supporting technologies, finds itself at a crucial crossroads, one where it is not merely extending its established routine roles within the realm of biomedicine. Instead, it is transforming into an invaluable tool that plays an increasingly significant and direct role in delivering cutting-edge, biomedically relevant information. The significant advancements that have recently been made in the field of mass spectrometry have notably broadened its prospective developmental directions and horizons, paving the way for novel explorations and applications. Looking forward to the future development of this transformative technology in the context of biomedicine, one can anticipate a variety of future perspectives and emerging trends that are on the horizon. Such trends could play a critical role in shaping its applications, particularly those associated with medical diagnostics, while also enhancing the overall value derived from patient-derived samples. The astonishing potential for future developments encompasses not just advancements in the technologies and methodologies that boost current analytical capabilities but also successful translational initiatives. These initiatives are expected to culminate in the return of information that is stratified based on patient data regarding various disease states and overall health conditions. This will present data with unprecedented amplitude and remarkably fast turn-around times, greatly surpassing the limitations imposed by the current standards in the field. Furthermore, the complex landscapes of genomics, proteomics, and metabolomics harbor intricately layered information that often emerges in isomeric or isobaric forms. This complexity showcases an extensive plethora of molecular species and their respective fates, which can range from various post-translational modifications (PTMs) and folding alterations to diverse forms of adducts and beyond. As a result, the continual development of specialized instruments alongside tailored methodologies that are proficient at addressing and unraveling this intricate array of information is absolutely essential for both biologists and clinicians. It is through these effective applications that the full potential of these advancements can be realized, ultimately allowing for greater exploitation of these tools and their deeper integration into routine clinical practice, fostering an environment where diagnostics can evolve and adapt to better serve patient needs. [23, 1, 201, 8, 67, 203, 303, 202, 10, 2]

The successful introduction of ambient ion sources serves as a clear and compelling demonstration of how meticulous, careful, and well-thought-out developments in the realm of advanced analytical technology can indeed open up entirely new avenues for exploration, as well as fresh perspectives on applications that were once deemed to be unthinkable and beyond reach with older instrumental configurations. It is widely foreseen that there will be a consequent and significant expansion in the numerical growth of this innovative and cutting-edge type of interface, and there exists no doubt regarding their remarkable ability to profoundly influence and impact the ever-evolving field of modern medicine in a very substantial and meaningful manner. This innovative method represents not only its very first instance but also stands out as one of the most formidable and challenging hurdles yet to be encountered in this rapidly evolving field. This challenge is particularly daunting, given that it requires a unique and specialized set of skills alongside an in-depth and comprehensive knowledge pertaining to thorough risk assessment; competencies which are traditionally not part of the standard expertise or training that scientists deeply immersed in biomolecular analysis usually possess. Over the last decade, mass spectrometry, along with its vast range of applications, has experienced a significant and continuous surge in growth and advancement across many disciplines. The ongoing and relentless improvement of sophisticated instruments alongside the development of associated methodologies has led to an unceasing broadening of the overall application scope in various diverse areas. This notable expansion has culminated in an increasingly widespread and routine application of mass spectrometry across numerous fields such as pharmacokinetics, metabolomics, and peptidomics, in addition to the more traditional area known as proteomics, which is primarily focused on the detailed and comprehensive study and analysis of pathological tissues. However, it soon became evidently clear that this advanced technology, in its classical set-up, was not truly suited to facilitate a full and comprehensive scientific appraisal and evaluation of the complex phenomena that involve alterations at the protein level within various intricate biological systems. This emerging body of knowledge ultimately resulted in the establishment of an entirely new scientific discipline, namely toxicoproteomics, whose main and primary aim is to apply the proteomic approach systematically in the thorough study and evaluation of adverse effects that are induced by xenobiotics and other foreign substances entering the intricate biological milieu, thereby deepening our understanding of how such compounds interact with biological systems at a fundamental level. [37, 236, 74, 55, 304, 305, 306, 307]

Chapter - 11

Conclusion

In conclusion, the recent exemplary development of new ionization mode coupled with mass spectrometry or of a new family of instruments has been evidence of the great attractions of mass spectrometry in biomedicine. It should not only play a more important role in this field but also a more direct one. It is envisioned to have a rapid diagnostic test and therapeutic monitoring device molded in every form factor present in our daily life. The advance of diagnostic medical devices will significantly drive the realization of this concept. As the crucial part of the device, mass spectrometry (MS) examination was also rapidly developed, particularly with the emergence of ambient ionization technology and incorporation with mini serial instruments. Such developments could change the paradigm from possible or probable to more certain diagnosis, and on provide critical information ensuring the success of personalized therapy. This trend is illustrated by the rotation to a more measurement-driven discovery of biomarkers and understanding the evolution of other "-omics" besides genomics and proteomics. Technologies relevant to omics are progressing at different rates. However, increasing numbers of positive results illustrate that they are ready for prime-time use on translational research. Understanding and management of MS-derived omics data are worries for broad acceptance of the findings.

The attention is moved to moving urine-based data towards practical implementation. This is conceptualized in a more structural and practical way as a pipeline of intelligence able to span the spectrum from data collection to knowledge systems. Although many of the concepts and technologies promoted in this volume require further research and validation, others are more readily assimilated. The attempt is made to build a coalescent strategy from those that are, embracing top-down and bottom-up development cosmetics. The progression of this work seems to timely coincide with a raised level of interest in urinary biomarkers as diagnostic, screening, epidemiological or risk biomarkers among commercial firms, regulatory bodies, clinicians, patients, authorized bodies, and the general public. There are generally a close link from the bench to bedside; the successful translation of urine-based discovery needs to be simultaneously addressed by protection of IP and continuous dialogue with key stakeholders. Chance of this translation should in part also drive the most suitable discovery and validation efforts topics. The aspiration is to leave a legacy that will serve as a path to furthering future research and as a framework for broad collaboration among researchers and also with an interest in new discovery in the clinical environment. This implies, amongst other things, an enhanced effort to shape the professional community and engaged society as the ultimate judge of malting the right decisions on omic technologies. It is considered it is important that all sought of professionals, patients and individuals are well informed on what this technology is (and is not) and its potential risks and benefits. Concomitant with this is there is a need to ensure the offering of a venue as open as possible for debate and a lively exchange of views on this subject. Despite the requirement in investment, instrumentation or personnel, all these aspects are easier to control. Implementing consistent practices will help to reconcile a myriad of variables and ensure the external validity of the trials. By contrast, other aspects are the unfamiliar and daunting territory for clinical and research teams undertaking omic profiling. This volume aims to guide investigators through the tangled web presented by bioinformatics, cheminformatics, and clinical investigations as well as the depths of complexity of large-scale data very rough guidelineed practices bar the results of numerous discussions with recognized experts and trial and error processes has been formulated.

References

1. A. G. Birhanu, "Mass spectrometry-based proteomics as an emerging tool in clinical laboratories," Clinical proteomics, 2023. springer.com
2. D. Li, J. Yi, G. Han, and L. Qiao, "MALDI-TOF mass spectrometry in clinical analysis and research," ACS measurement science au, 2022. acs.org
3. X. Zhan, J. Li, Y. Guo, and O. Golubnitschaja, "Mass spectrometry analysis of human tear fluid biomarkers specific for ocular and systemic diseases in the context of 3P medicine," EPMA Journal, 2021. springer.com
4. I. Jalaludin, D. M. Lubman, and J. Kim, "A guide to mass spectrometric analysis of extracellular vesicle proteins for biomarker discovery," Mass Spectrometry Reviews, 2023. [HTML]
5. E. Ponzini, C. Santambrogio, A. De Palma, *et al.*, "Mass spectrometry-based tear proteomics for noninvasive biomarker discovery," Mass Spectrometry, vol. 2022, Wiley Online Library. wiley.com
6. A. Pulumati, A. Pulumati, B. S. Dwarakanath, "Technological advancements in cancer diagnostics: Improvements and limitations," *Cancer*, vol. 2023, Wiley Online Library. wiley.com
7. Y. Hu, Z. Wang, L. Liu, J. Zhu, D. Zhang, M. Xu, *et al.*, "Mass spectrometry-based chemical mapping and profiling toward molecular understanding of diseases in precision medicine," *Chemical*, vol. 2021. rsc.org
8. V. G. Zaikin and R. S. Borisov, "Mass spectrometry as a crucial analytical basis for omics sciences," Journal of Analytical Chemistry, 2021. springer.com
9. N. M. Lima, B. L. M. Fernandes, G. F. Alves, J. C. Q. de Souza, *et al.*, "Mass spectrometry applied to diagnosis, prognosis, and therapeutic targets identification for the novel coronavirus SARS-CoV-2: A review," *Analytica Chimica Acta*, vol. 1196, p. 339-356, 2022. nih.gov
10. C. B. Messner, V. Demichev, Z. Wang, J. Hartl, *et al.*, "Mass spectrometry-based high-throughput proteomics and its role in biomedical studies and systems biology," 2023. wiley.com

11. H. M. Britt, T. Cragnolini, and K. Thalassinos, "Integration of mass spectrometry data for structural biology," Chemical reviews, 2021. ucl.ac.uk

12. S. Yan, R. Bhawal, Z. Yin, T. W. Thannhauser, "Recent advances in proteomics and metabolomics in plants," Molecular Horticulture, vol. 2, no. 1, pp. 1-12, 2022. springer.com

13. C. Zhang, J. Xu, R. Tang, J. Yang, W. Wang, and X. Yu, "Novel research and future prospects of artificial intelligence in cancer diagnosis and treatment," *Journal of Hematology & Oncology*, vol. 16, no. 1, 2023. springer.com

14. L. Gozdzialski, B. Wallace, and D. Hore, "Point-of-care community drug checking technologies: an insider look at the scientific principles and practical considerations," Harm Reduction Journal, 2023. springer.com

15. G. Visconti, J. Boccard, M. Feinberg, and S. Rudaz, "From fundamentals in calibration to modern methodologies: a tutorial for small molecules quantification in liquid chromatography–mass spectrometry …," Analytica chimica acta, 2023. sciencedirect.com

16. Z. Wang, H. Zhu, and W. Xiong, "Advances in mass spectrometry-based multi-scale metabolomic methodologies and their applications in biological and clinical investigations," Science Bulletin, 2023. sciencedirect.com

17. D. Yadav, "Analytical Chemistry Foundations," 2025. [HTML]

18. A. K. Kuril, "Exploring the versatility of mass spectrometry: Applications across diverse scientific disciplines," European Journal of Mass Spectrometry, 2024. [HTML]

19. T. Fujii, "Exploring Modern Mass Spectrometry and Its Real-World Applications," Springer, . [HTML]

20. I. Siddique, "High-performance liquid chromatography: comprehensive techniques and cutting-edge innovations," Journal of Advances in Engineering and Technology, 2023. ssrn.com

21. T. Guo, J. A. Steen, and M. Mann, "Mass-spectrometry-based proteomics: from single cells to clinical applications," Nature, 2025. [HTML]

22. P. Madhiyazhagan, "Advances in Instrumentation for Biological Sciences: Unveiling the Secrets of Life," Advance Research Trends in Biology, 2024. researchgate.net

23. W. Zhang, X. Wang, Y. Xia, and Z. Ouyang, "Ambient Ionization and Miniature Mass Spectrometry Systems for Disease Diagnosis and Therapeutic Monitoring," 2017. ncbi.nlm.nih.gov
24. X. Chen, W. Shu, L. Zhao, and J. Wan, "Advanced mass spectrometric and spectroscopic methods coupled with machine learning for in vitro diagnosis," View, 2023. wiley.com
25. K. Kratka, P. Sistik, I. Olivkova, "Mass Spectrometry–Based Proteomics in Clinical Diagnosis of Amyloidosis and Multiple Myeloma: A Review (2012–2024)," *Mass Spectrometry*, vol. 2025, Wiley Online Library. wiley.com
26. E. Deschamps, V. Calabrese, I. Schmitz, M. Hubert-Roux, "Advances in ultra-high-resolution mass spectrometry for pharmaceutical analysis," Molecules, 2023. mdpi.com
27. J. Zhang, M. Sans, K. Y. Garza, *et al*., "Mass spectrometry technologies to advance care for cancer patients in clinical and intraoperative use," Mass Spectrometry, vol. 2021, Wiley Online Library. [HTML]
28. G. Cutshaw, S. Uthaman, N. Hassan, *et al*., "The emerging role of Raman spectroscopy as an omics approach for metabolic profiling and biomarker detection toward precision medicine," Chemical, vol. 2023, ACS Publications. nih.gov
29. H. Hu and J. Laskin, "Emerging computational methods in mass spectrometry imaging," Advanced Science, 2022. wiley.com
30. A. Isa, M. Gharibi, A. Cetinkaya, and S. A. Ozkan, "Sustainable and scalable detection: Paper-based analytical devices and miniaturized detection systems for modern diagnostics," Microchemical Journal, 2025. [HTML]
31. J. R. Bonney and B. M. Prentice, "Perspective on emerging mass spectrometry technologies for comprehensive lipid structural elucidation," Analytical chemistry, 2021. nih.gov
32. S. R. Shuken, "An introduction to mass spectrometry-based proteomics," Journal of proteome research, 2023. stevenshuken.com
33. B. Greer, O. Chevallier, B. Quinn, and L. M. Botana, "... and shoot: The evolution of the technique and its application in the analysis of foods and biological matrices by liquid chromatography mass spectrometry," TrAC Trends in Analytical Chemistry, vol. 143, 2021. sciencedirect.com
34. A. Bauermeister and H. Mannochio-Russo, "Mass spectrometry-based

metabolomics in microbiome investigations," *Nature Reviews*, 2022. nih.gov

35. T. Van Acker, S. Theiner, E. Bolea-Fernandez, *et al.*, "Inductively coupled plasma mass spectrometry," *Nature Reviews*, 2023. [HTML]
36. W. Bahureksa, M. M. Tfaily, R. M. Boiteau, "Soil organic matter characterization by fourier transform ion cyclotron resonance mass spectrometry (FTICR MS): A critical review of sample preparation, analysis, and …," ACS Publications, 2021. [HTML]
37. M. J. Taylor and J. K. Lukowski, "Spatially resolved mass spectrometry at the single cell: recent innovations in proteomics and metabolomics," *Journal of the American Society for Mass Spectrometry*, vol. 32, no. 1, pp. 1-10, 2021. acs.org
38. Y. Brachthäuser, "Development and characterization of a Fourier Transform based Quadrupole Ion Trap (FT-QIT) technique for process and residual gas analysis," 2018. [PDF]
39. K. Arturi and J. Hollender, "Machine learning-based hazard-driven prioritization of features in nontarget screening of environmental high-resolution mass spectrometry data," Environmental Science & Technology, 2023. acs.org
40. H. Yue, F. He, Z. Zhao, and Y. Duan, "Plasma-based ambient mass spectrometry: Recent progress and applications," Mass Spectrometry Reviews, 2023. [HTML]
41. A. Kulyyassov, M. Fresnais, and R. Longuespée, "Targeted liquid chromatography-tandem mass spectrometry analysis of proteins: Basic principles, applications, and perspectives," Proteomics, 2021. [HTML]
42. E. S. Nakayasu, M. Gritsenko, P. D. Piehowski, Y. Gao, *et al.*, "Tutorial: best practices and considerations for mass-spectrometry-based protein biomarker discovery and validation," *Nature*, vol. 2021. nature.com
43. Y. H. Lai and Y. S. Wang, "Advances in high-resolution mass spectrometry techniques for analysis of high mass-to-charge ions," Mass Spectrometry Reviews, 2023. [HTML]
44. P. Babele and A. K. Yadav, "Back2Basics: mass-to-charge ratio (m/z) in proteomics," Journal of Proteins and Proteomics, 2023. [HTML]
45. M. F. Jarrold, "Applications of charge detection mass spectrometry in molecular biology and biotechnology," Chemical reviews, 2021. nih.gov

46. E. Desligničre, A. Rolland, E. H. T. M. Ebberink, *et al.*, "Orbitrap-based mass and charge analysis of single molecules," *Accounts of Chemical Research*, vol. 2023, ACS Publications. acs.org
47. L. M. Miller and M. F. Jarrold, "Charge detection mass spectrometry for the analysis of viruses and virus-like particles," Essays in biochemistry, 2023. nih.gov
48. A. Özdemir, J. L. Lin, M. Gülfen, and C. H. Chen, "Simultaneous detection of positively and negatively charged microparticles by ion trap mass spectrometry," Analytical Methods, 2022. [HTML]
49. S. H. Lai, S. Tamara, and A. J. R. Heck, "Single-particle mass analysis of intact ribosomes by mass photometry and Orbitrap-based charge detection mass spectrometry," Iscience, 2021. cell.com
50. I. A. Kaltashov, D. G. Ivanov, and Y. Yang, "Mass spectrometry-based methods to characterize highly heterogeneous biopharmaceuticals, vaccines, and nonbiological complex drugs at the intact-mass level," Mass spectrometry reviews, 2024. nih.gov
51. G. R. D. Prabhu, E. R. Williams, M. Wilm, "Mass spectrometry using electrospray ionization," *Nature Reviews Methods*, vol. 2023. [HTML]
52. K. C. Fabijanczuk, J. W. Hager, and S. A. McLuckey, "Separation and Simultaneous Trapping of Multiply Charged and Singly Charged Ions for Mass Spectrometry: Application to Lipid Mixtures," Analytical chemistry, 2023. nih.gov
53. W. Wang, Y. Chen, L. Li, L. Zhou, X. Du, and M. Liu, "Chemical composition of different size ultrafine particulate matter measured by nanoparticle chemical ionization mass spectrometer," *Journal of Environmental*, vol. 2022, Elsevier. [HTML]
54. S. Heiles, "Advanced tandem mass spectrometry in metabolomics and lipidomics-methods and applications," Analytical and Bioanalytical Chemistry, 2021. springer.com
55. H. M. Bennett, W. Stephenson, C. M. Rose, and S. Darmanis, "Single-cell proteomics enabled by next-generation sequencing or mass spectrometry," Nature Methods, 2023. [HTML]
56. T. Züllig and H. C. Köfeler, "High resolution mass spectrometry in lipidomics," Mass spectrometry reviews, 2021. wiley.com
57. S. Alseekh, A. Aharoni, Y. Brotman, K. Contrepois, *et al.*, "Mass spectrometry-based metabolomics: a guide for annotation, quantification and best reporting practices," *Nature Methods*, 2021. nih.gov

58. E. L. Boys, J. Liu, P. J. Robinson, and R. R. Reddel, "Clinical applications of mass spectrometry-based proteomics in cancer: where are we?," Proteomics, 2023. wiley.com

59. R. Noberini, G. Robusti, and T. Bonaldi, "Mass spectrometry-based characterization of histones in clinical samples: applications, progress, and challenges," The FEBS Journal, 2022. wiley.com

60. Y. J. Jung, H. S. Seo, J. H. Kim, K. Y. Song, C. H. Park, "… of volatile organic compounds real time analysis analysis from exhaled breath of gastric cancer patients using proton-transfer-reaction time-of-flight mass spectrometry," Frontiers in …, 2021. frontiersin.org

61. M. Korecka and L. M. Shaw, "Mass spectrometry-based methods for robust measurement of Alzheimer's disease biomarkers in biological fluids," Journal of neurochemistry, 2021. wiley.com

62. A. Khoo, L. Y. Liu, J. O. Nyalwidhe, O. J. Semmes, "Proteomic discovery of non-invasive biomarkers of localized prostate cancer using mass spectrometry," *Nature Reviews*, 2021. nih.gov

63. C. T. McDowell, X. Lu, A. S. Mehta, *et al.*, "Applications and continued evolution of glycan imaging mass spectrometry," Mass spectrometry, 2023. nih.gov

64. J. E. Kim, D. Han, J. S. Jeong, J. J. Moon, H. K. Moon, *et al.*, "Multisample mass spectrometry-based approach for discovering injury markers in chronic kidney disease," Molecular & Cellular, vol. XX, no. YY, pp. ZZ-ZZ, 2021. sciencedirect.com

65. J. G. Flitcroft, J. Verheyen, T. Vemulkar, *et al.*, "Early detection of kidney cancer using urinary proteins: a truly non-invasive strategy," BJU International, vol. 2022. [HTML]

66. S. Mondal and S. Banerjee, "Desorption Electrospray Ionization Mass Spectrometry Imaging for Cancer Diagnosis," in *… and Biophysical Methods in Molecular and …*, Springer, 2025. [HTML]

67. J. Hu, F. Liu, Y. Chen, G. Shangguan *et al.*, "Mass spectrometric biosensing: a powerful approach for multiplexed analysis of clinical biomolecules," ACS sensors, 2021. nju.edu.cn

68. A. Vinaiphat, J. K. Low, K. W. Yeoh, W. J. Chng, "Application of advanced mass spectrometry-based proteomics to study hypoxia driven cancer progression," *Frontiers in …*, 2021. frontiersin.org

69. L. Shi, A. Habib, L. Bi, H. Hong, R. Begum, "Ambient ionization mass spectrometry: application and prospective," Critical Reviews in…, vol. 2024, Taylor & Francis. researchgate.net

70. S. Dash, C. C. Wu, C. C. Wu, S. F. Chiang, Y. T. Lu, and C. Y. Yeh, "… vesicle membrane protein profiling and targeted mass spectrometry unveil CD59 and tetraspanin 9 as novel plasma biomarkers for detection of colorectal cancer," Cancers, vol. 14, no. 1, 2022. mdpi.com

71. X. Ma and F. M. Fernández, "Advances in mass spectrometry imaging for spatial cancer metabolomics," Mass spectrometry reviews, 2024. nih.gov

72. J. Paul and T. D. Veenstra, "Separation of serum and plasma proteins for in-depth proteomic analysis," Separations, 2022. mdpi.com

73. C. Li, S. Chu, S. Tan, X. Yin, Y. Jiang, and X. Dai, "Towards higher sensitivity of mass spectrometry: A perspective from the mass analyzers," Frontiers in…, 2021. frontiersin.org

74. S. Tamara, M. A. den Boer, and A. J. R. Heck, "High-resolution native mass spectrometry," Chemical reviews, 2021. acs.org

75. T. J. Bruno, J. W. Robinson, and E. M. S. Frame, "Mass Spectrometry I: Principles and Instrumentation," *Undergraduate …*, 2023. [HTML]

76. Z. Yang, Z. Ren, Y. Cheng, W. Sun, Z. Xi, W. Jia, and G. Li, "Review and prospect on portable mass spectrometer for recent applications," *Vacuum*, vol. 2022, Elsevier. [HTML]

77. R. E. March and F. A. Londry, "Theory of quadrupole mass spectrometry," in *Practical Aspects of Ion Trap Mass Spectrometry*, 2024. [HTML]

78. K. O. Nagornov, O. Y. Tsybin, and E. Nicol, "Fourier transform ion cyclotron resonance mass spectrometry at the true cyclotron frequency," Mass Spectrometry, vol. 2022, Wiley Online Library. [HTML]

79. T. Tucholski and Y. Ge, "Fourier-transform ion cyclotron resonance mass spectrometry for characterizing proteoforms," Mass spectrometry reviews, 2022. nih.gov

80. Y. Qi, P. Fu, and D. A. Volmer, "Analysis of natural organic matter via fourier transform ion cyclotron resonance mass spectrometry: an overview of recent non-petroleum applications," Mass Spectrometry Reviews, 2022. wiley.com

81. C. Feng, S. Liu, T. Jiang, and W. Xu, "Multiplexing Quadrupole and Ion Trap Operation Modes on a “Brick” Miniature Mass Spectrometer," Molecules, 2023. mdpi.com
82. X. Fang, J. Xie, S. Chu, Y. Jiang, Y. An, C. Li, and X. Gong, "Quadrupole-linear ion trap tandem mass spectrometry system for clinical biomarker analysis," *Engineering*, vol. 2022, Elsevier. sciencedirect.com
83. J. S. Bhanot, K. C. Fabijanczuk, A. M. Abdillahi, *et al*., "Adaptation and operation of a quadrupole/time-of-flight tandem mass spectrometer for high mass ion/ion reaction studies," *Journal of Mass Spectrometry*, vol. 2022, Elsevier. sciencedirect.com
84. C. Li, S. Chu, S. Tan, X. Yin *et al*., "Towards Higher Sensitivity of Mass Spectrometry: A Perspective From the Mass Analyzers," 2021. ncbi.nlm.nih.gov
85. E. Zarrouk, M. Lenski, C. Bruno, V. Thibert, "High-resolution mass spectrometry: Theoretical and technological aspects," *Analytique et Clinique*, 2022. sciencedirect.com
86. T. Dickel and A. Mollaebrahimi, "Unveiling nuclear isomers through multiple-reflection time-of-flight mass spectrometry," The European Physical Journal Special Topics, 2024. springer.com
87. A. Mathew, R. Buijs, G. B. Eijkel, F. Giskes, *et al*., "Ion imaging of native protein complexes using orthogonal time-of-flight mass spectrometry and a Timepix detector," *Journal of Mass Spectrometry*, vol. 2021, ACS Publications. acs.org
88. K. Vashisht, P. Sethi, and K. Sharma, "Basic Principles, Instrumentation and Applications of Mass Spectrometry," in *Guidebook on Spectroscopic …*, 2024. [HTML]
89. S. Bräkling, "Development of a Dual Ionization Source Time-of-Flight Mass Spectrometer for Gas Chromatography," 2023. uni-wuppertal.de
90. S. Perchepied, Z. Zhou, G. Mitulović, "Exploiting ion-mobility mass spectrometry for unraveling proteome complexity," *Journal of Separation*, vol. 2023, Wiley Online Library. vub.be
91. A. Romano, V. Capozzi, I. Khomenko, and F. Biasioli, "Advances in the application of direct injection mass spectrometry techniques to the analysis of grape, wine and other alcoholic beverages," Molecules, 2023. mdpi.com

92. A. Fayyaz, M. A. Baig, M. Waqas, and U. Liaqat, "… -Induced Breakdown Spectroscopy (LIBS), MFA-LIBS, Thermal LIBS, Laser Ablation Time-of-Flight Mass Spectrometry, Energy-Dispersive X-ray Spectroscopy …," Minerals, 2024. mdpi.com

93. W. Jeong, H. S. Kim, and B. Kang, "Development of a portable time-of-flight mass spectrometer prototype using a cold electron source," Review of Scientific Instruments, 2022. [HTML]

94. D. G. Delafield, G. Lu, C. J. Kaminsky, and L. Li, "High-end ion mobility mass spectrometry: A current review of analytical capacity in omics applications and structural investigations," TrAC Trends in Analytical Chemistry, vol. 2022, Elsevier. sciencedirect.com

95. A. Kiontke, S. Billig, and C. Birkemeyer, "Response in Ambient Low Temperature Plasma Ionization Compared to Electrospray and Atmospheric Pressure Chemical Ionization for Mass Spectrometry," 2018. ncbi.nlm.nih.gov

96. A. Henderson and L. M. Heaney, "Ambient ionisation mass spectrometry for drug and toxin analysis: A review of the recent literature," *Testing and Analysis*, vol. 2024, Wiley Online Library. wiley.com

97. K. M. Shimko, J. W. O'Brien, J. Li, B. J. Tscharke, "In-sewer stability assessment of anabolic steroids and selective androgen receptor modulators," *Environmental Science & Technology*, vol. 2022, ACS Publications. unit.no

98. S. A. Borden, "Paper Spray Mass Spectrometry for the Quantitation of Drugs of Abuse in Biofluids and Street Drug Samples," 2023. uvic.ca

99. J. Wen, L. Duan, B. Wang, Q. Dong, Y. Liu, C. Chen, "In-sewer stability assessment of 140 pharmaceuticals, personal care products, pesticides and their metabolites: Implications for wastewater-based …", Environment, 2024. sciencedirect.com

100. H. N. Mtetwa, I. D. Amoah, S. Kumari, F. Bux *et al.*, "Optimisation of analytical methods for tuberculosis drug detection in wastewater: A multinational study," Heliyon, 2024. cell.com

101. R. Aslan, D. Y. Ovat, A. Atasoy Aydin, "A 7-year study of workplace drug testing in two major cities in Türkiye," Journal of Ethnicity in ..., vol. XX, no. XX, pp. XX-XX, 2024. ankara.edu.tr

102. F. Frankenfeld, L. Wagmann, *et al.*, "Studies on the Stability and Microbial Biotransformation of Five Deschloroketamine Derivatives as

Prerequisite for Wastewater-Based Epidemiology Screening," *Testing and Analysis*, 2024. wiley.com

103. C. Aydoğan, B. Çakan, S. Alharthi, A. Ali, I. Göktürk, "Recent advances and applications in drug analysis by nano-scale separations techniques," Green Analytical Chemistry, vol. 2024, Elsevier. sciencedirect.com

104. A. I. Al-Asmari, "A critical review of workplace drug testing methods for old and new psychoactive substances: Gaps, advances, and perspectives," Saudi Pharmaceutical Journal, 2024. sciencedirect.com

105. B. H. Lapizco-Encinas, Y. V. Zhang, P. P. Gqamana, *et al.*, "Capillary electrophoresis as a sample separation step to mass spectrometry analysis: a primer," TrAC Trends in Analytical Chemistry, vol. XX, no. YY, pp. ZZ-ZZ, 2023. sciencedirect.com

106. M. R. Savina, B. H. Isselhardt, and R. Trappitsch, "Simultaneous isotopic analysis of U, Pu, and Am in spent nuclear fuel by resonance ionization mass spectrometry," Analytical Chemistry, 2021. osti.gov

107. S. M. Marshall, C. Mathis, E. Carrick, and G. Keenan, "Identifying molecules as biosignatures with assembly theory and mass spectrometry," *Nature*, 2021. nature.com

108. T. R. Hawkinson and R. C. Sun, "Matrix-assisted laser desorption/ionization mass spectrometry imaging of glycogen in situ," in *Mass Spectrometry Imaging of Small Molecules*, Springer, 2021. [HTML]

109. D. Unsihuay and D. Mesa Sanchez, "Quantitative mass spectrometry imaging of biological systems," *Annual Review of Physical Chemistry*, vol. 72, pp. 1-26, 2021. nih.gov

110. J. Hermann, L. Schurgers, and V. Jankowski, "Identification and characterization of post-translational modifications: Clinical implications," Molecular aspects of medicine, 2022. sciencedirect.com

111. Y. Zhao, Q. Xue, M. Wang, B. Meng, and Y. Jiang, "Evolution of mass spectrometry instruments and techniques for blood proteomics," *Journal of Proteome*, vol. 2023, ACS Publications. [HTML]

112. Q. Fu, C. I. Murray, O. A. Karpov, *et al.*, "Automated proteomic sample preparation: The key component for high throughput and quantitative mass spectrometry analysis," Mass Spectrometry, 2023. nih.gov

113. A. Pandey, "Mass Spectrometry: Techniques and Applications," 2025. [HTML]

114. M. L. Spruill, M. Maletic-Savatic, H. Martin, F. Li, *et al.*, "Spatial analysis of drug absorption, distribution, metabolism, and toxicology using mass spectrometry imaging," *Biochemical Journal*, vol. 2022, Elsevier. sciencedirect.com

115. B. C. Orsburn, Y. Yuan, and N. N. Bumpus, "Insights into protein post-translational modification landscapes of individual human cells by trapped ion mobility time-of-flight mass spectrometry," Nature communications, 2022. nature.com

116. C. D. Gutierrez-Reyes, P. Jiang, M. Atashi, *et al.*, "Advances in mass spectrometry-based glycoproteomics: an update covering the period 2017–2021," Wiley Online Library, 2022. wiley.com

117. A. Pandey, "Mass Spectrometry: Techniques and Applications," 2025. [HTML]

118. L. Lamont, D. Hadavi, B. Viehmann, B. Flinders, "Quantitative mass spectrometry imaging of drugs and metabolites: a multiplatform comparison," *Analytical and Bioanalytical Chemistry*, vol. 413, no. 1, pp. 123-134, 2021. springer.com

119. H. H. Maurer, "Hyphenated high-resolution mass spectrometry-the "all-in-one" device in analytical toxicology?," Analytical and Bioanalytical Chemistry, 2021. springer.com

120. R. Dziewas and E. Michou, "European Stroke Organisation and European Society for Swallowing Disorders guideline for the diagnosis and treatment of post-stroke dysphagia," *European Stroke Journal*, vol. 6, no. 1, pp. 1-16, 2021. sagepub.com

121. N. Gabrielly Pereira dos Santos, E. Vasconcelos Soares Maciel, D. Arley Vargas Medina, and F. Mauro Lanças, "NanoLC-EI-MS: Perspectives in Biochemical Analysis," 2023. ncbi.nlm.nih.gov

122. E. Christofi and P. Barran, "Ion mobility mass spectrometry (IM-MS) for structural biology: insights gained by measuring mass, charge, and collision cross section," Chemical reviews, 2023. acs.org

123. A. Kim, P. F. Kelly, M. A. Turner, "A direct analysis method using sheath flow probe electrospray ionisation-mass spectrometry (sfPESI-MS) to detect drug residues from fingerprint forensic gel lifts," Drug Testing and Analysis, 2025. wiley.com

124. M. Grabarics, M. Lettow, C. Kirschbaum, K. Greis, "Mass spectrometry-based techniques to elucidate the sugar code," Chemical Reviews, vol. 121, no. 1, pp. 123-145, 2021. acs.org

125. L. Piersimoni, P. L. Kastritis, C. Arlt, and A. Sinz, "Cross-linking mass spectrometry for investigating protein conformations and protein–protein Interactions─ A method for all seasons," Chemical Reviews, 2021. acs.org

126. F. Raposo and D. Barceló, "Challenges and strategies of matrix effects using chromatography-mass spectrometry: An overview from research versus regulatory viewpoints," TrAC Trends in Analytical Chemistry, 2021. [HTML]

127. H. , H. F. Wu, and S. K. Kailasa, "Recent advances in nanomaterials-based optical sensors for detection of various biomarkers (inorganic species, organic and biomolecules)," Luminescence, 2023. [HTML]

128. A. Ahmad, M. Imran, and H. Ahsan, "Biomarkers as biomedical bioindicators: Approaches and techniques for the detection, analysis, and validation of novel biomarkers of diseases," Pharmaceutics, 2023. mdpi.com

129. L. Xu, M. Ma, J. Li, D. Dai, D. Gao, P. Ma, and Q. Wu, "Exploration of aminopeptidase N as new biomarker for early diagnosis of thyroid cancer," *Biosensors and...,* vol. 2024, Elsevier. jlu.edu.cn

130. D. L. Ng, A. C. Granados, Y. A. Santos, V. Servellita, *et al.*, "A diagnostic host response biosignature for COVID-19 from RNA profiling of nasal swabs and blood," *Science*, vol. 2021. science.org

131. D. Yadav, K. K. Pvsn, S. Tomo, S. Sankanagoudar, "Association of iron-related biomarkers with severity and mortality in COVID-19 patients," Journal of Trace Elements in Medicine and Biology, vol. XX, no. YY, pp. ZZ-ZZ, 2022. nih.gov

132. O. M. Peiro, A. Carrasquer, R. Sanchez-Gimenez, and others, "Biomarkers and short-term prognosis in COVID-19," Biomarkers, vol. 26, no. 1, pp. 1-10, 2021. nih.gov

133. S. Rasheed, T. Kanwal, N. Ahmad, B. Fatima, "Advances and challenges in portable optical biosensors for onsite detection and point-of-care diagnostics," TrAC Trends in Analytical Chemistry, vol. 2024, Elsevier. [HTML]

134. D. D. Vallejo, C. Rojas Ramírez, K. F. Parson, Y. Han, "Mass spectrometry methods for measuring protein stability," Chemical Reviews, vol. 2022, ACS Publications. nih.gov

135. R. Hu, Y. Li, Y. Yang, and M. Liu, "Mass spectrometry-based strategies for single-cell metabolomics," Mass Spectrometry Reviews, 2023. [HTML]

136. A. McKenzie-Coe, N. S. Montes, and L. M. Jones, "Hydroxyl radical protein footprinting: a mass spectrometry-based structural method for studying the higher order structure of proteins," Chemical reviews, 2021. acs.org

137. C. C. Eckhoff, N. K. Lubinsky, L. J. Metzler, *et al*., "Low-Cost, Compact Quadrupole Mass Filters with Unity Mass Resolution via Ceramic Resin Vat Photopolymerization," *Advanced*, 2024. wiley.com

138. L. Yang, Y. Gu, H. Gong, Z. Liu *et al*., "Development and research of the quadrupole mass spectrometry simulation model with the entire ion optics system," Scientific Reports, 2025. nature.com

139. E. Defossez, J. Bourquin, S. von Reuss, "Eight key rules for successful data-dependent acquisition in mass spectrometry-based metabolomics," Mass Spectrometry, vol. 2023, Wiley Online Library. wiley.com

140. M. Bergin, D. J. Ward, S. M. Lambrick, and others, "Low-energy electron ionization mass spectrometer for efficient detection of low mass species," *Review of Scientific Instruments*, vol. 2021. aip.org

141. J. Ahamad, F. Ali, M. A. Sayed, and J. Ahmad, "Basic principles and fundamental aspects of mass spectrometry," in *Mass Spectrometry in...*, 2022. [HTML]

142. A. G. Nikolaev, E. M. Oks, V. P. Frolova, K. P. Savkin, "Time-of-Flight Analyzer of the Plasma Mass-to-Charge Composition with a Secondary Electronic Multiplier," gdpa.ru. gdpa.ru

143. A. A. Reinhard, J. D. Inglis, M. J. Singleton, A. H. Kara, "Application of ATONA amplifiers to the measurement of uranium isotopic ratios by thermal ionization mass spectrometry," *Analytical Chemistry*, vol. 2024, ACS Publications. [HTML]

144. N. A. Zirakparvar, B. Manard, C. Hexel, D. Dunlap, "Faraday detector uranium isotope ratio measurement: Insights from solution-and laser ablation-based sampling methodologies on the Neoma MC-ICP-MS," *Journal of Mass Spectrometry*, 2023. sciencedirect.com

145. Z. He, W. Wang, L. Zhai, Z. Li, H. Deng, T. Feng, "A gated integration method for the detection of unevenly distributed ions bunches in laser resonant ionization mass spectrometry," *Journal of Mass Spectrometry*, vol. 2021, Elsevier. [HTML]

146. P. Boutachkov, K. O. Voss, K. Lee, M. S. Song, C. Yi, "An investigation of secondary electron emission from ZnO based nanomaterials for future applications in radiation detectors," *Scientific Reports*, 2021. nature.com

147. E. Bolli, A. Bellucci, M. Mastellone, A. Mezzi, "Engineered SnO2-based thin films for efficient CO2 gas sensing at room temperature," *Applied Surface Science*, vol. 2025, Elsevier. sciencedirect.com

148. L. Poirier, A. Bayerle, A. Lassise, F. Torretti, R. Schupp, and others, "Cross-calibration of a combined electrostatic and time-of-flight analyzer for energy-and charge-state-resolved spectrometry of tin laser-produced plasma," *Applied Physics B*, vol. 128, no. 3, 2022. springer.com

149. N. Dontschuk, L. V. H. Rodgers, J. P. Chou, "X-ray quantification of oxygen groups on diamond surfaces for quantum applications," *Materials for Quantum Applications*, 2023. iop.org

150. A. Vanella, "Nanoparticle formation in nanoporous structures and applications," 2022. unisi.it

151. A. Belousov, M. Miller, R. Continetti, "Sampling accelerated micron scale ice particles with a quadrupole ion trap mass spectrometer," *Journal of the …*, vol. 2021, ACS Publications. researchgate.net

152. C. Zhang, R. Terabayashi, and S. Hasegawa, "Design and simulation for a 90Sr ion trap–laser cooling spectrometer," Nuclear Instruments and Methods, 2025. sciencedirect.com

153. J. Gasser, A. Galli, and P. Wurz, "Absolute beam monitor: A novel laboratory device for neutral beam calibration," Review of scientific instruments, 2022. aip.org

154. A. A. K. Cramer, "Design and applications of cold-cathode X-ray imaging systems," 2021. mit.edu

155. J. N. Audinot, P. Philipp, O. De Castro, *et al.*, "Highest resolution chemical imaging based on secondary ion mass spectrometry performed on the helium ion microscope," Reports on Progress, vol. 2021. [HTML]

156. N. Alygizakis, F. Lestremau, P. Gago-Ferrero, *et al.*, "Towards a harmonized identification scoring system in LC-HRMS/MS based non-target screening (NTS) of emerging contaminants," TrAC Trends in Analytical Chemistry, vol. 2023, Elsevier. google.com

157. P. Bonetti, C. Leuz, and G. Michelon, "Large-sample evidence on the

impact of unconventional oil and gas development on surface waters," Science, 2021. science.org

158. A. Merchant, S. Batzner, S. S. Schoenholz, M. Aykol, *et al.*, "Scaling deep learning for materials discovery," Nature, 2023. nature.com

159. S. N. Toala, Z. Sun, Y. Yue, S. F. Gonski *et al.*, "Recent developments in ionophore-based potentiometric electrochemical sensors for oceanic carbonate detection," Sensors & Diagnostics, 2024. rsc.org

160. F. Wyler, M. A. Oestreich, B. S. Frauchiger, *et al.*, "Correction of sensor crosstalk error in Exhalyzer D multiple-breath washout device significantly impacts outcomes in children with cystic fibrosis," *Journal of Applied Physiology*, vol. 2021. physiology.org

161. J. Lou, Y. Jiao, R. Yang, Y. Huang, X. Xu, and L. Zhang, "Calibration-free, high-precision, and robust terahertz ultrafast metasurfaces for monitoring gastric cancers," *Proceedings of the National Academy of Sciences*, 2022. pnas.org

162. L. Cole, M. Clench, and S. Francese, "Sample treatment for tissue proteomics in cancer, toxicology, and forensics," 2019. [PDF]

163. L. Gatto, R. Aebersold, J. Cox, V. Demichev, J. Derks, *et al.*, "Initial recommendations for performing, benchmarking and reporting single-cell proteomics experiments," *Nature*, 2023. nih.gov

164. C. J. Chen, D. Y. Lee, J. Yu, Y. N. Lin, *et al.*, "Recent advances in LC-MS-based metabolomics for clinical biomarker discovery," Mass Spectrometry, vol. 2023, Wiley Online Library. [HTML]

165. A. Delvaux and E. Rathahao-Paris, "Different ion mobility-mass spectrometry coupling techniques to promote metabolomics," Mass Spectrometry, vol. 2022, Wiley Online Library. [HTML]

166. S. Lee and K. Chintalapudi, "Clinical chemistry for developing countries: mass spectrometry," *Annual Review of …*, vol. 2021. nih.gov

167. V. P. Gaspar, S. Ibrahim, R. P. Zahedi, "Utility, promise, and limitations of liquid chromatography-mass spectrometry-based therapeutic drug monitoring in precision medicine," *Journal of Mass Spectrometry*, vol. 2021, Wiley Online Library. wiley.com

168. B. A. Rappold, "Review of the use of liquid chromatography-tandem mass spectrometry in clinical laboratories: part I-development," Annals of Laboratory Medicine, 2022. koreamed.org

169. X. Zhou, W. Zhang, and Z. Ouyang, "Recent advances in on-site mass spectrometry analysis for clinical applications," TrAC Trends in Analytical Chemistry, 2022. nih.gov

170. J. Therriault, M. S. Woo, G. Salvadó, J. Gobom, and others, "Comparison of immunoassay-with mass spectrometry-derived p-tau quantification for the detection of Alzheimer's disease pathology," Molecular, vol. 2024, Springer. springer.com

171. S. Rankin-Turner and L. M. Heaney, "Mass spectrometry in the clinical laboratory. A short journey through the contribution to the scientific literature by CCLM," Clinical Chemistry and Laboratory Medicine, vol. 2023. degruyter.com

172. V. C. Wasinger, M. Zeng, and Y. Yau, "Current Status and Advances in Quantitative Proteomic Mass Spectrometry," 2013. ncbi.nlm.nih.gov

173. M. Nikolov, C. Schmidt, and H. Urlaub, "Quantitative mass spectrometry-based proteomics: An overview.," 2012. [PDF]

174. NPM Smit, LR Ruhaak, FP Romijn, "The time has come for quantitative protein mass spectrometry tests that target unmet clinical needs," *Journal of Mass Spectrometry*, vol. 2021, ACS Publications. acs.org

175. M. E. King, M. Lin, M. Spradlin, "Advances and emerging medical applications of direct mass spectrometry technologies for tissue analysis," *Annual Review of Analytical Chemistry*, vol. 2023. annualreviews.org

176. B. A. Rappold, "Review of the use of liquid chromatography-tandem mass spectrometry in clinical laboratories: Part II-Operations," Ann Lab Med, 2022. koreamed.org

177. W. L. Cheng, C. Markus, C. Y. Lim, R. Z. Tan, "Calibration practices in clinical mass spectrometry: Review and recommendations," Annals of Laboratory Medicine, 2023. koreamed.org

178. R. A. Groves, M. Mapar, R. Aburashed, L. F. Ponce, *et al.*, "Methods for quantifying the metabolic boundary fluxes of cell cultures in large cohorts by high-resolution hydrophilic liquid chromatography mass spectrometry," *ACS Chemistry*, vol. 2022. acs.org

179. S. Lingxiao, M. Jinfei, Y. Qingqing, L. Pengyun *et al.*, "The application of certified reference materials for clinical mass spectrometry," 2022. ncbi.nlm.nih.gov

180. H. P. Wang, P. Chen, J. W. Dai, D. Liu, J. Y. Li, and Y. P. Xu, "Recent

advances of chemometric calibration methods in modern spectroscopy: Algorithms, strategy, and related issues," *Trends in Analytical Chemistry*, vol. 2022, Elsevier. [HTML]

181. K. Mervič, V. S. Šelih, M. Šala, "Non-matrix-matched calibration in bulk multi-element laser ablation–inductively coupled plasma–mass spectrometry analysis of diverse materials," 2024. ssrn.com

182. J. A. Kirwan, H. Gika, R. D. Beger, D. Bearden, W. B. Dunn, *et al.*, "Quality assurance and quality control reporting in untargeted metabolic phenotyping: mQACC recommendations for analytical quality management," Metabolomics, vol. 18, no. 12, 2022. springer.com

183. S. Li, J. Chen, W. Xu, B. Sun, J. Wu, Q. Chen, "Highly homogeneous bimetallic core–shell Au@ Ag nanoparticles with embedded internal standard fabrication using a microreactor for reliable quantitative SERS," *Materials Chemistry*, vol. 2023. [HTML]

184. A. Nasiri, R. Jahani, S. Mokhtari, and H. Yazdanpanah, "Overview, consequences, and strategies for overcoming matrix effects in LC-MS analysis: a critical review," Analyst, vol. 146, no. 1, pp. 1-15, 2021. [HTML]

185. M. Lafuente, S. De Marchi, M. Urbiztondo, *et al.*, "Plasmonic MOF thin films with Raman internal standard for fast and ultrasensitive SERS detection of chemical warfare agents in ambient air," *ACS Publications*, 2021. acs.org

186. H. Zhao, Q. Lin, L. Huang, Y. Zhai, Y. Liu, Y. Deng, and E. Su, "… immunoassay with enhanced precision for the detection of cTnI amplified by acridinium ester-loaded microspheres and internally calibrated by magnetic fluorescent …," Nanoscale, 2021. [HTML]

187. G. Xie, L. Wang, T. Chen, K. Zhou, Z. Zhang, and J. Li, "A metabolite array technology for precision medicine," *Analytical Chemistry*, vol. 93, no. xx, pp. xx-xx, 2021. [HTML]

188. K. A. Lippa, J. J. Aristizabal-Henao, R. D. Beger, J. A. Bowden, *et al.*, "Reference materials for MS-based untargeted metabolomics and lipidomics: a review by the metabolomics quality assurance and quality control consortium (mQACC)," Metabolomics, vol. 18, no. 1, pp. 1-15, 2022. springer.com

189. I. H. T. Guideline, "Validation of analytical procedures Q2 (R2)," ICH: Geneva, . citeline.com

190. J. M. Ruijter, R. J. Barnewall, I. B. Marsh, *et al.*, "Efficiency correction is required for accurate quantitative PCR analysis and reporting," *Clinical Chemistry*, vol. 2021. oup.com

191. H. Joy Finley-Jones, "Improving figures of merit and expanding applications for inductively coupled plasma mass spectrometry," 2010. [PDF]

192. A. Katal, S. Dahiya, and T. Choudhury, "Energy efficiency in cloud computing data centers: a survey on software technologies," Cluster Computing, 2023. springer.com

193. S. Kumar, T. Gopi, N. Harikeerthana, and M. K. Gupta, "Machine learning techniques in additive manufacturing: a state of the art review on design, processes and production control," Journal of Intelligent Manufacturing, vol. 2023, Springer. springer.com

194. J. A. Zwerts, P. J. Stephenson, F. Maisels, *et al*., "Methods for wildlife monitoring in tropical forests: Comparing human observations, camera traps, and passive acoustic sensors," *Conservation Science and Practice*, vol. 2021, Wiley Online Library. wiley.com

195. W. Wahyudi, "Five components of work motivation in the achievement of lecturer performance," Scientific Journal of Reflection: Economic, Accounting, vol. XX, no. YY, pp. ZZ-ZZ, 2022. ojspustek.org

196. H. M. Alzoubi, "The effect of electronic human resources management on organizational health of telecommunications companies in Jordan," *Journal of Data and Network Science*, vol. 2022. skylineuniversity.ac.ae

197. M. Adaikkappan *et al*., "Modeling, state of charge estimation, and charging of lithium-ion battery in electric vehicle: a review," *International Journal of …*, 2022. wiley.com

198. M. R. Costa-Jussà, J. Cross, O. Çelebi, M. Elbayad, "No language left behind: Scaling human-centered machine translation," arXiv preprint arXiv:XXXX.XXXX, 2022. [PDF]

199. D. J. Deming, "Four facts about human capital," Journal of Economic Perspectives, 2022. aeaweb.org

200. Z. Usmani, M. Sharma, A. K. Awasthi, and T. Lukk, "Lignocellulosic biorefineries: the current state of challenges and strategies for efficient commercialization," in *Renewable and Sustainable Energy Reviews*, vol. 135, 2021. aston.ac.uk

201. J. L. Moore, N. H. Patterson, J. L. Norris, "Prospective on imaging mass spectrometry in clinical diagnostics," Molecular & Cellular, vol. 2023, Elsevier. sciencedirect.com

202. H. Lee and S. I. Kim, "Review of liquid chromatography-mass spectrometry-based proteomic analyses of body fluids to diagnose infectious diseases," International Journal of Molecular Sciences, 2022. mdpi.com

203. A. J. Grooms and B. J. Burris, "Mass spectrometry for metabolomics analysis: Applications in neonatal and cancer screening," Mass Spectrometry, 2024. nih.gov

204. F. A. Castelli, G. Rosati, C. Moguet, C. Fuentes, "Metabolomics for personalized medicine: the input of analytical chemistry from biomarker discovery to point-of-care tests," *Analytical and Bioanalytical Chemistry*, vol. 414, no. 1, pp. 1-10, 2022. springer.com

205. A. Pandey, "Mass Spectrometry: Techniques and Applications," 2025. [HTML]

206. J. Aoki and M. Toyoda, "Development of novel projection-type imaging mass spectrometer," Review of Scientific Instruments, 2021. [HTML]

207. Y. Li, T. Kind, J. Folz, A. Vaniya *et al*., "Spectral entropy outperforms MS/MS dot product similarity for small-molecule compound identification," Nature Methods, 2021. nih.gov

208. T. Baba, P. Ryumin, E. Duchoslav, K. Chen, "Dissociation of biomolecules by an intense low-energy electron beam in a high sensitivity time-of-flight mass spectrometer," *Society for Mass Spectrometry*, 2021. [HTML]

209. R. Bushuiev, A. Bushuiev, N. de Jonge, "Massspecgym: A benchmark for the discovery and identification of molecules," in Advances in ..., 2024. neurips.cc

210. L. R. Heil, E. Damoc, T. N. Arrey, A. Pashkova, *et al*., "Evaluating the performance of the astral mass analyzer for quantitative proteomics using data-independent acquisition," *Journal of Proteome Research*, vol. 2023, ACS Publications. acs.org

211. J. Koopman and S. Grimme, "From QCEIMS to QCxMS: A tool to routinely calculate CID mass spectra using molecular dynamics," *Journal of the American Society for Mass Spectrometry*, vol. 32, no. 10, pp. 2021, 2021. [HTML]

212. K. Scheubert, F. Hufsky, and S. Böcker, "Computational mass spectrometry for small molecules," 2013. ncbi.nlm.nih.gov

213. H. S. Nogay and H. Adeli, "Diagnostic of autism spectrum disorder based on structural brain MRI images using, grid search optimization, and convolutional neural networks," Biomedical Signal Processing and Control, 2023. uludag.edu.tr

214. Y. Perez-Riverol, J. Bai, C. Bandla, *et al.*, "The PRIDE database resources in 2022: a hub for mass spectrometry-based proteomics evidences," *Nucleic Acids Research*, vol. 50, no. D1, pp. D543-D552, 2022. oup.com

215. A. J. Miles, R. W. Janes, and B. A. Wallace, "Tools and methods for circular dichroism spectroscopy of proteins: a tutorial review," Chemical Society Reviews, 2021. rsc.org

216. M. A. Hoffmann, L. F. Nothias, M. Ludwig, *et al.*, "High-confidence structural annotation of metabolites absent from spectral libraries," *Nature*, vol. 2022. nature.com

217. Z. Pang, Y. Lu, G. Zhou, F. Hui, L. Xu, C. Viau, *et al.*, "MetaboAnalyst 6.0: towards a unified platform for metabolomics data processing, analysis and interpretation," *Nucleic Acids Research*, 2024. oup.com

218. M. Varadi and S. Velankar, "The impact of AlphaFold Protein Structure Database on the fields of life sciences," Proteomics, 2023. wiley.com

219. L. B. Ayres, F. J. V. Gomez, J. R. Linton, and M. F. Silva, "Taking the leap between analytical chemistry and artificial intelligence: A tutorial review," Analytica Chimica Acta, vol. 2021, Elsevier. [HTML]

220. J. A. Blakeley-Ruiz and M. Kleiner, "Considerations for constructing a protein sequence database for metaproteomics," *Computational and Structural Biotechnology Journal*, vol. 20, pp. 123-135, 2022. sciencedirect.com

221. O. I. Talantseva, R. S. Romanova, E. M. Shurdova, *et al.*, "The global prevalence of autism spectrum disorder: A three-level meta-analysis," Frontiers in…, 2023. frontiersin.org

222. M. Bai, R. Tang, G. Li, W. She, G. Chen, H. Shen, and S. Zhu, "High-throughput screening of 756 chemical contaminants in aquaculture products using liquid chromatography/quadrupole time-of-flight mass spectrometry," Food Chemistry: X, vol. 2022, Elsevier. sciencedirect.com

223. J. S. Brodbelt, "Deciphering combinatorial post-translational modifications by top-down mass spectrometry," Current opinion in chemical biology, 2022. sciencedirect.com

224. K. B. Donohoo, J. Wang, M. Goli, A. Yu, W. Peng, *et al.*, "Advances in mass spectrometry-based glycomics-An update covering the period 2017–2021," Wiley Online Library, 2022. wiley.com

225. S. A. Jarmusch, J. J. J. van der Hooft, P. C. Dorrestein, *et al.*, "Advancements in capturing and mining mass spectrometry data are transforming natural products research," Natural Product Reports, vol. 2021. nih.gov

226. R. L. Zhu and E. Jonas, "Rapid approximate subset-based spectra prediction for electron ionization–mass spectrometry," Analytical chemistry, 2023. acs.org

227. S. G. Lias, "Numeric Databases for Chemical Analysis," 1989. ncbi.nlm.nih.gov

228. A. Tripathi, Y. Vázquez-Baeza, J. M. Gauglitz, *et al.*, "Chemically informed analyses of metabolomics mass spectrometry data with Qemistree," *Nature Chemical Biology*, vol. 17, no. 3, pp. 322-331, 2021. nih.gov

229. M. Khalikova, J. Jireš, O. Horáček, "What is the role of current mass spectrometry in pharmaceutical analysis?" Mass Spectrometry, 2024. wiley.com

230. A. Pandey, "Mass Spectrometry: Techniques and Applications," 2025. [HTML]

231. Y. Liu, T. De Vijlder, W. Bittremieux, "Current and future deep learning algorithms for tandem mass spectrometry (MS/MS)-based small molecule structure elucidation," *Mass Spectrometry*, 2021. uantwerpen.be

232. S. Castellino and N. M. Lareau, "The emergence of imaging mass spectrometry in drug discovery and development: making a difference by driving decision making," *Journal of Mass Spectrometry*, vol. 2021, Wiley Online Library. wiley.com

233. A. Fiorante, L. A. Ye, A. Tata, T. Kiyota, M. Woolman, "A workflow for meaningful interpretation of classification results from handheld ambient mass spectrometry analysis probes," *International Journal of …*, 2024. mdpi.com

234. Y. Nowatzky, F. F. Russo, J. Lisec, A. Kister, *et al.*, "FIORA: Local neighborhood-based prediction of compound mass spectra from single fragmentation events," *Nature*, 2025. nature.com

235. N. M. Lima, D. V. A. de Aguiar, and B. G. Vaz, "Harnessing Mass Spectrometry for Molecular Analysis in Energy: From Fossil Fuels to Renewables," Annual Review of Analytical Chemistry, vol. 2025. [HTML]

236. F. Meissner, J. Geddes-McAlister, M. Mann, "The emerging role of mass spectrometry-based proteomics in drug discovery," *Nature Reviews Drug Discovery*, vol. 21, no. 5, pp. 345-367, 2022. [HTML]

237. Y. Qi, Q. Xie, J. J. Wang, D. He, H. Bao, Q. L. Fu, S. Su, "Deciphering dissolved organic matter by Fourier transform ion cyclotron resonance mass spectrometry (FT-ICR MS): from bulk to fractions and individuals," Carbon Research, vol. 2022, Springer. springer.com

238. J. Yuk Kong, "Development of Tandem Mass Spectrometric Methods and Instrumentation for the Structural Elucidation of Unknown Drug Metabolites Based on Ion/Molecule Reactions," 2018. [PDF]

239. R. Patel and A. Patel, "Advancements in high-resolution analytical techniques for pharmaceutical characterization: Bridging quality control and drug efficacy," *International Journal of Innovative Research in …*, 2024. researchgate.net

240. S. A. Tittlemier, B. Cramer, M. C. DeRosa, Z. Dzuman, *et al.*, "Developments in analytical techniques for mycotoxin determination: an update for 2022-23," World Mycotoxin Journal, vol. 17, no. 1, pp. 1-10, 2024. brill.com

241. A. Godana Birhanu, "Mass spectrometry-based proteomics as an emerging tool in clinical laboratories," 2023. ncbi.nlm.nih.gov

242. A. Alexandridou, P. Schorr, C. S. Stokes, *et al.*, "Analysis of vitamin D metabolic markers by mass spectrometry: Recent progress regarding the 'gold standard' method and integration into clinical practice," Mass Spectrometry, vol. 2023, Wiley Online Library. wiley.com

243. S. Tsuchida and T. Nakayama, "MALDI-based mass spectrometry in clinical testing: Focus on bacterial identification," Applied Sciences, 2022. mdpi.com

244. A. L. Marie, S. Ray, S. Lu, J. Jones, I. Ghiran, *et al.*, "High-sensitivity glycan profiling of blood-derived immunoglobulin G, plasma, and

extracellular vesicle isolates with capillary zone electrophoresis-mass spectrometry," *ACS Chemistry*, vol. 2021. nih.gov

245. M. Shanmuganathan, Z. Kroezen, B. Gill, S. Azab, *et al.*, "The maternal serum metabolome by multisegment injection-capillary electrophoresis-mass spectrometry: a high-throughput platform and standardized data workflow," *Nature*, vol. 2021. [HTML]

246. R. Lehmann, "… to bench-practical considerations to avoid pre-analytical pitfalls and assess sample quality for high-resolution metabolomics and lipidomics analyses of body fluids," Analytical and Bioanalytical Chemistry, 2021. springer.com

247. P. A. Doble, R. G. de Vega, D. P. Bishop, D. J. Hare, "Laser ablation–inductively coupled plasma–mass spectrometry imaging in biology," Chemical Reviews, vol. 2021, ACS Publications. uts.edu.au

248. Y. Liu, Z. Wu, D. W. Armstrong, H. Wolosker, *et al.*, "Detection and analysis of chiral molecules as disease biomarkers," *Nature Reviews*, vol. 2023. nih.gov

249. C. Ivarsson Orrelid, O. Rosberg, S. Weiner, "Applying machine learning to high-dimensional proteomics datasets for the identification of Alzheimer's disease biomarkers," Fluids and Barriers of …, vol. 2025, Springer. springer.com

250. G. Paglia, A. J. Smith, and G. Astarita, "Ion mobility mass spectrometry in the omics era: Challenges and opportunities for metabolomics and lipidomics," Mass spectrometry reviews, 2022. [HTML]

251. L. Mattoli, M. Gianni, and M. Burico, "Mass spectrometry-based metabolomic analysis as a tool for quality control of natural complex products," Mass Spectrometry Reviews, 2023. [HTML]

252. S. L. Collins, I. Koo, J. M. Peters, P. B. Smith, and others, "Current challenges and recent developments in mass spectrometry–based metabolomics," *Annual Review of...,* vol. 2021. [HTML]

253. C. Mas-Bargues, C. Escriva, M. Dromant, C. Borras, *et al.*, "Lipid peroxidation as measured by chromatographic determination of malondialdehyde. Human plasma reference values in health and disease," *Archives of Biochemistry*, vol. 2021, Elsevier. sciencedirect.com

254. JYW Ma, YH Sze, JF Bian, "Critical role of mass spectrometry proteomics in tear biomarker discovery for multifactorial ocular

diseases," *Journal of Molecular*, vol. 2021, pp. 1-10, 2021. spandidos-publications.com

255. F. Boel, V. Akimov, M. Teuchler, M. K. Terkelsen, *et al.*, "Deep proteome profiling of metabolic dysfunction-associated steatotic liver disease," *Communications*, 2025. nature.com

256. A. S. Marchev, L. V. Vasileva, K. M. Amirova, *et al.*, "Metabolomics and health: From nutritional crops and plant-based pharmaceuticals to profiling of human biofluids," *Cellular and Molecular*, vol. 2021, Springer. springer.com

257. A. Pandey, "Mass Spectrometry: Techniques and Applications," 2025. [HTML]

258. S. Gobena, B. Admassu, M. Z. Kinde, "Proteomics and its current application in biomedical area: Concise review," The Scientific World Journal, vol. 2024, Wiley Online Library. wiley.com

259. S. Asif, Y. Wenhui, S. ur-Rehman, Q. ul-ain, "Advancements and prospects of machine learning in medical diagnostics: unveiling the future of diagnostic precision," Methods in Engineering, 2024. [HTML]

260. D. Yadav, "Analytical Chemistry Foundations," 2025. [HTML]

261. P. Ahmad and A. Hussain, "Mass spectrometry-based proteomic approaches for salivary protein biomarkers discovery and dental caries diagnosis: A critical review," Mass Spectrometry, vol. 2024, Wiley Online Library. [HTML]

262. Y. Wang, R. Li, W. Shu, X. Chen *et al.*, "Designed Nanomaterials-Assisted Proteomics and Metabolomics Analysis for In Vitro Diagnosis," Small Methods, 2024. [HTML]

263. A. N. Neagu, M. Jayathirtha, E. Baxter, M. Donnelly, *et al.*, "Applications of tandem mass spectrometry (MS/MS) in protein analysis for biomedical research," Molecules, 2022. mdpi.com

264. M. T. Patel *et al.*, "Advancements in cancer diagnostics: integrating surface-enhanced Raman spectroscopy and microptofluidics for precision and versatility," *Applied Spectroscopy*, 2025. tandfonline.com

265. J. A. Paulo, V. Kadiyala, P. A. Banks, H. Steen *et al.*, "Mass Spectrometry-Based Proteomics for Translational Research: A Technical Overview," 2012. ncbi.nlm.nih.gov

266. N. Titkare, S. Chaturvedi, S. Borah, "Advances in mass spectrometry for metabolomics: Strategies, challenges, and innovations in disease biomarker discovery," Biomedical Engineering, vol. 2024, Wiley Online Library. [HTML]

267. V. S. K. Kumar, "REVIEW ON THE ROLE OF MASS SPECTROMETRY IN MODERN BIOANALYTICAL METHOD DEVELOPMENT," 2024. amazonaws.com

268. A. Son, W. Kim, J. Park, Y. Park, W. Lee, S. Lee, *et al.*, "Mass Spectrometry Advancements and Applications for Biomarker Discovery, Diagnostic Innovations, and Personalized Medicine," *Journal of Molecular Sciences*, 2024. mdpi.com

269. K. K. Krestensen, R. M. A. Heeren, and B. Balluff, "State-of-the-art mass spectrometry imaging applications in biomedical research," Analyst, 2023. [HTML]

270. T. Meisinger, A. Vogt, R. Kretz, and H. S. Hammer, "Mass spectrometry-based ligand binding assays in biomedical research," Expert Review of..., 2025. Taylor & Francis. [HTML]

271. G. Mitra, "FULL LENGTH PAPER_ Evolving role of mass spectrometry in modern biomedical research and biopharmaceuticals," NHBT, 2021. hithaldia.ac.in

272. C. Lin, Q. Tian, S. Guo, D. Xie *et al.*, "Metabolomics for Clinical Biomarker Discovery and Therapeutic Target Identification," 2024. ncbi.nlm.nih.gov

273. W. Bittremieux, M. Wang, and P. C. Dorrestein, "The critical role that spectral libraries play in capturing the metabolomics community knowledge," Metabolomics, 2022. nih.gov

274. S. Shen, C. Zhan, C. Yang, A. R. Fernie *et al.*, "Metabolomics-centered mining of plant metabolic diversity and function: Past decade and future perspectives," Molecular plant, 2023. cell.com

275. B. C. Muthubharathi, T. Gowripriya, and K. Balamurugan, "Metabolomics: small molecules that matter more," Molecular omics, 2021. [HTML]

276. J. Zhou and L. Zhong, "Applications of liquid chromatography-mass spectrometry based metabolomics in predictive and personalized medicine," Frontiers in Molecular Biosciences, 2022. frontiersin.org

277. R. Katam, C. Lin, K. Grant, C. S. Katam, and S. Chen, "Advances in

plant metabolomics and its applications in stress and single-cell biology," *International Journal of …*, 2022. mdpi.com

278. D. Feng, H. Li, T. Xu, F. Zheng *et al.*, "… metabolomics and cellular heterogeneity exploration by inertial microfluidics coupled with pulsed electric field-induced electrospray ionization-high resolution mass …," Analytica Chimica Acta, 2022. [HTML]

279. S. Guo, S. Qiu, Y. Cai, Z. Wang, Q. Yang, and S. Tang, "Mass spectrometry-based metabolomics for discovering active ingredients and exploring action mechanism of herbal medicine," *Frontiers in …*, 2023. frontiersin.org

280. S. P. Putri, M. M. M. Ikram, A. Sato, H. A. Dahlan, "Application of gas chromatography-mass spectrometry-based metabolomics in food science and technology," *Journal of Bioscience*, vol. 2022, Elsevier. [HTML]

281. L. Li, J. Han, Z. Wang, J. Liu *et al.*, "Mass Spectrometry Methodology in Lipid Analysis," 2014. ncbi.nlm.nih.gov

282. T. Xu, C. Hu, Q. Xuan, and G. Xu, "Recent advances in analytical strategies for mass spectrometry-based lipidomics," 2020. ncbi.nlm.nih.gov

283. R. Matthiesen, C. Lauber, J. L. Sampaio, N. Domingues, *et al.*, "Shotgun mass spectrometry-based lipid profiling identifies and distinguishes between chronic inflammatory diseases," *The Lancet*, vol. XX, no. YY, pp. ZZ-ZZ, 2021. thelancet.com

284. T. G. Meikle, K. Huynh, C. Giles, and P. J. Meikle, "Clinical lipidomics: realizing the potential of lipid profiling," Journal of Lipid Research, 2021. sciencedirect.com

285. V. Solntceva and M. Kostrzewa, "Detection of species-specific lipids by routine MALDI TOF mass spectrometry to unlock the challenges of microbial identification and antimicrobial susceptibility testing," *Frontiers in Cellular and Infection Microbiology*, vol. 11, 2021. frontiersin.org

286. L. Katz, A. Tata, M. Woolman, and A. Zarrine-Afsar, "Lipid profiling in cancer diagnosis with hand-held ambient mass spectrometry probes: addressing the late-stage performance concerns," Metabolites, 2021. mdpi.com

287. R. Lerner, D. Baker, C. Schwitter, S. Neuhaus, "Four-dimensional

trapped ion mobility spectrometry lipidomics for high throughput clinical profiling of human blood samples," *Nature*, 2023. nature.com

288. H. Zhang, D. G. Delafield, and L. Li, "Mass spectrometry imaging: the rise of spatially resolved single-cell omics," Nature methods, 2023. researchgate.net

289. A. Mund, A. D. Brunner, and M. Mann, "Unbiased spatial proteomics with single-cell resolution in tissues," Molecular cell, 2022. cell.com

290. C. J. Good, E. K. Neumann, C. E. Butrico, J. E. Cassat, "High spatial resolution MALDI imaging mass spectrometry of fresh-frozen bone," *Analytical Chemistry*, vol. 2022, ACS Publications. nih.gov

291. YR Xie, DC Castro, SS Rubakhin, TJ Trinklein, "Multiscale biochemical mapping of the brain through deep-learning-enhanced high-throughput mass spectrometry," Nature, vol. 2024. nature.com

292. D. C. Koomen, J. C. May, and J. A. McLean, "Insights and prospects for ion mobility-mass spectrometry in clinical chemistry," Expert review of proteomics, 2022. nih.gov

293. T. Xu, Q. Wang, Q. Wang, and L. Sun, "Mass spectrometry-intensive top-down proteomics: an update on technology advancements and biomedical applications," Analytical Methods, 2024. rsc.org

294. A. M. L. Santilli, K. Ren, R. Oleschuk, "Application of intraoperative mass spectrometry and data analytics for oncological margin detection, a review," IEEE Transactions, vol. 2022. [HTML]

295. M. N. Stillger, M. J. Li, P. Hönscheid, C. von Neubeck, and others, "Advancing rare cancer research by MALDI mass spectrometry imaging: Applications, challenges, and future perspectives in sarcoma," Wiley Online Library, 2024. wiley.com

296. G. Gradisteanu Pircalabioru and M. Raileanu, "Fast detection of bacterial gut pathogens on miniaturized devices: an overview," *Diagnostics*, vol. 2024, Taylor & Francis. tandfonline.com

297. L. Guo, Y. Wang, Z. Liu, F. Zhang *et al*., "EQLC-EC: An Efficient Voting Classifier for 1D Mass Spectrometry Data Classification," Electronics, 2025. mdpi.com

298. K. Hořejší, D. Kolářová, R. Jirásko, and M. Holčapek, "Recent advances, challenges, and future directions in the mass spectrometry analysis of glycosphingolipids in biological samples," TrAC Trends in Analytical Chemistry, vol. 2024, Elsevier. sciencedirect.com

299. K. M. Engel, P. Prabutzki, J. Leopold, A. Nimptsch, *et al.*, "A new update of MALDI-TOF mass spectrometry in lipid research," *Progress in Lipid Research*, vol. 2022, Elsevier. [HTML]

300. C. Harkin, K. W. Smith, F. L. Cruickshank, *et al.*, "On-tissue chemical derivatization in mass spectrometry imaging," Mass Spectrometry, vol. 2022, Wiley Online Library. wiley.com

301. G. Joseph, B. Binny, and A. R. Venter, "Nanobubbles in Electrospray Ionization Mass Spectrometry," Analytical Chemistry, 2025. acs.org

302. K. R. Karch, D. T. Snyder, S. R. Harvey, "Native mass spectrometry: recent progress and remaining challenges," Annual Review of ..., vol. 2022. nih.gov

303. A. Pandey, "Mass Spectrometry: Techniques and Applications," 2025. [HTML]

304. J. A. Melby, D. S. Roberts, E. J. Larson, *et al.*, "Novel strategies to address the challenges in top-down proteomics," *Mass Spectrometry*, vol. 2021, ACS Publications. nih.gov

305. L. M. Grobbelaar, C. Venter, M. Vlok, M. Ngoepe, *et al.*, "SARS-CoV-2 spike protein S1 induces fibrin (ogen) resistant to fibrinolysis: implications for microclot formation in COVID-19," *Bioscience*, 2021. portlandpress.com

306. M. Schirmer and C. Dusny, "Microbial single-cell mass spectrometry: status, challenges, and prospects," Current Opinion in Biotechnology, 2023. [HTML]

307. D. Smith, P. Španěl, N. Demarais, "Recent developments and applications of selected ion flow tube mass spectrometry (SIFT-MS)," Mass Spectrometry, vol. 2025, Wiley Online Library. wiley.com